

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101923\
 Data File : VY016040.D
 Acq On : 19 Oct 2023 16:22
 Operator : SY/MD
 Sample : 04938-06
 Misc : 3.80g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A-3(2-6)

Quant Time: Oct 20 05:38:13 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 19 03:15:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	140801	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	236862	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	212931	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	86353	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	88116	55.062	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.120%
35) Dibromofluoromethane	7.716	113	78291	50.217	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.440%
50) Toluene-d8	10.179	98	280335	49.509	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.020%
62) 4-Bromofluorobenzene	12.483	95	84767	41.956	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	83.920%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101923\
Data File : VY016040.D
Acq On : 19 Oct 2023 16:22
Operator : SY/MD
Sample : 04938-06
Misc : 3.80g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A-3(2-6)

Quant Time: Oct 20 05:38:13 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 19 03:15:44 2023
Response via : Initial Calibration

