

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101923\
 Data File : VY016042.D
 Acq On : 19 Oct 2023 17:09
 Operator : SY/MD
 Sample : 04938-12
 Misc : 4.24g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 A-6(2-6)

Quant Time: Oct 20 05:38:52 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 19 03:15:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	139739	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	234997	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	216698	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	100301	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	87437	55.053	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	110.100%
35) Dibromofluoromethane	7.716	113	77120	49.858	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	99.720%
50) Toluene-d8	10.179	98	278260	49.533	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.060%
62) 4-Bromofluorobenzene	12.483	95	93258	46.525	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	93.060%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.710	84	11322	6.473	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101923\
Data File : VY016042.D
Acq On : 19 Oct 2023 17:09
Operator : SY/MD
Sample : 04938-12
Misc : 4.24g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
A-6(2-6)

Quant Time: Oct 20 05:38:52 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 19 03:15:44 2023
Response via : Initial Calibration

