

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101923\
 Data File : VY016051.D
 Acq On : 19 Oct 2023 20:40
 Operator : SY/MD
 Sample : 04843-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 WC-1

Quant Time: Oct 20 05:42:07 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 19 03:15:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	42827	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	67700	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	56392	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	23337	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	22383	45.984	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	91.960%
35) Dibromofluoromethane	7.722	113	21794	48.908	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.820%
50) Toluene-d8	10.185	98	77314	47.772	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.540%
62) 4-Bromofluorobenzene	12.489	95	22329	38.667	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	77.340%
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.710	84	3997	7.457	ug/l #	74

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY101923\
Data File : VY016051.D
Acq On : 19 Oct 2023 20:40
Operator : SY/MD
Sample : 04843-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
WC-1

Quant Time: Oct 20 05:42:07 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 19 03:15:44 2023
Response via : Initial Calibration

