

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102022\
 Data File : VY011021.D
 Acq On : 20 Oct 2022 13:42
 Operator : KP/MD
 Sample : N5193-01
 Misc : 3.10g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LSP-SS-01-00-20

Quant Time: Oct 20 22:22:58 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	334145	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	571606	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	471582	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	185355	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	198017	64.334	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	128.660%
35) Dibromofluoromethane	7.716	113	201582	63.863	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	127.720%
50) Toluene-d8	10.179	98	719161	60.256	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	120.520%
62) 4-Bromofluorobenzene	12.477	95	218499	51.792	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	103.580%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102022\
Data File : VY011021.D
Acq On : 20 Oct 2022 13:42
Operator : KP/MD
Sample : N5193-01
Misc : 3.10g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LSP-SS-01-00-20

Quant Time: Oct 20 22:22:58 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

