

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102022\
 Data File : VY011022.D
 Acq On : 20 Oct 2022 14:05
 Operator : KP/MD
 Sample : N5193-02
 Misc : 6.09g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LSP-DS-01-09-11

Quant Time: Oct 20 22:23:07 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	301014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	499578	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	390193	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	124726	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	165149	59.561	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 119.120%	
35) Dibromofluoromethane	7.716	113	160676	58.045	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 116.080%	
50) Toluene-d8	10.179	98	602749	57.783	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 115.560%	
62) 4-Bromofluorobenzene	12.483	95	165214	44.366	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 88.740%	
Target Compounds						Qvalue
16) Acetone	3.948	43	17898	22.228	ug/l	90
17) Carbon Disulfide	4.192	76	10211	1.220	ug/l	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102022\
Data File : VY011022.D
Acq On : 20 Oct 2022 14:05
Operator : KP/MD
Sample : N5193-02
Misc : 6.09g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LSP-DS-01-09-11

Quant Time: Oct 20 22:23:07 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration

