

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102022\
 Data File : VY011029.D
 Acq On : 20 Oct 2022 16:43
 Operator : KP/MD
 Sample : N5193-12
 Misc : 5.88g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 LSP-SS-05-08-13

Quant Time: Oct 20 22:24:47 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
 Quant Title : SW846 8260
 QLast Update : Tue Oct 18 01:42:42 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) Pentafluorobenzene	7.789	168	278519	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	468543	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	336739	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	99858	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	167620	65.334	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 130.660%	
35) Dibromofluoromethane	7.716	113	170402	65.930	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 131.860%	
50) Toluene-d8	10.179	98	553303	56.557	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 113.120%	
62) 4-Bromofluorobenzene	12.477	95	142950	40.677	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 81.360%	
Target Compounds						
					Qvalue	
16) Acetone	3.942	43	55182	74.066	ug/l	# 89
25) 2-Butanone	6.978	43	24618	21.789	ug/l	95
84) 1,2,4-Trimethylbenzene	13.117	105	8592	1.439	ug/l	96
95) Naphthalene	15.233	128	58197	15.791	ug/l	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102022\
Data File : VY011029.D
Acq On : 20 Oct 2022 16:43
Operator : KP/MD
Sample : N5193-12
Misc : 5.88g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
LSP-SS-05-08-13

Quant Time: Oct 20 22:24:47 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101722S.M
Quant Title : SW846 8260
QLast Update : Tue Oct 18 01:42:42 2022
Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.789 min Scan# 91
Delta R.T. -0.000 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43




Instrument :
MSVOA_Y
ClientSampleId :
LSP-SS-05-08-13


Tgt Ion:168 Resp: 278519
Ion Ratio Lower Upper
168 100
99 54.1 46.9 70.3



#16
Acetone
Concen: 74.066 ug/l
RT: 3.942 min Scan# 348
Delta R.T. -0.000 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43

Tgt Ion: 43 Resp: 55182
Ion Ratio Lower Upper
43 100
58 33.4 22.0 33.0#









#25
2-Butanone
Concen: 21.789 ug/l
RT: 6.978 min Scan# 846
Delta R.T. 0.000 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43




Instrument :
MSVOA_Y
ClientSampleId :
LSP-SS-05-08-13


Tgt Ion: 43 Resp: 24618
Ion Ratio Lower Upper
43 100
72 24.7 17.7 26.5



#33
1,2-Dichloroethane-d4
Concen: 65.334 ug/l
RT: 8.136 min Scan# 1036
Delta R.T. 0.000 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43

Tgt Ion: 65 Resp: 167620
Ion Ratio Lower Upper
65 100
67 54.3 0.0 99.4





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.685 min Scan# 1126

Delta R.T. 0.000 min

Lab File: VY011029.D

Acq: 20 Oct 2022 16:43

Instrument :

MSVOA_Y

ClientSampleId :

LSP-SS-05-08-13





Tgt Ion:114 Resp: 468543

Ion Ratio Lower Upper

114 100

63 19.0 0.0 43.2

88 15.5 0.0 31.8



#35

Dibromofluoromethane

Concen: 65.930 ug/l

RT: 7.716 min Scan# 967

Delta R.T. 0.006 min

Lab File: VY011029.D

Acq: 20 Oct 2022 16:43




Tgt Ion:113 Resp: 170402


Ion Ratio Lower Upper

113 100

111 101.9 82.6 124.0

192 19.3 17.6 26.4









#50
Toluene-d8
Concen: 56.557 ug/l
RT: 10.179 min Scan# 11
Delta R.T. -0.000 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43




Instrument :
MSVOA_Y
ClientSampleId :
LSP-SS-05-08-13


Tgt Ion: 98 Resp: 553303
Ion Ratio Lower Upper
98 100
100 65.4 50.2 75.2



#62
4-Bromofluorobenzene
Concen: 40.677 ug/l
RT: 12.477 min Scan# 1748
Delta R.T. -0.006 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43

Tgt Ion: 95 Resp: 142950
Ion Ratio Lower Upper
95 100
174 83.1 0.0 177.0
176 80.2 0.0 175.8









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.489 min Scan# 11
Delta R.T. -0.001 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43




Instrument :
MSVOA_Y
ClientSampleId :
LSP-SS-05-08-13


Tgt Ion:117 Resp: 336739
Ion Ratio Lower Upper
117 100
82 54.1 43.0 64.4
119 32.3 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.422 min Scan# 1903
Delta R.T. -0.000 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43

Tgt Ion:152 Resp: 99858
Ion Ratio Lower Upper
152 100
115 57.6 28.2 84.7
150 157.8 0.0 347.6









#84
1,2,4-Trimethylbenzene
Concen: 1.439 ug/l
RT: 13.117 min Scan# 1853
Delta R.T. 0.000 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43

Instrument :
MSVOA_Y
ClientSampleId :
LSP-SS-05-08-13

Tgt Ion:105 Resp: 8592
Ion Ratio Lower Upper
105 100
120 47.2 22.3 66.9



#95
Naphthalene
Concen: 15.791 ug/l
RT: 15.233 min Scan# 2200
Delta R.T. -0.000 min
Lab File: VY011029.D
Acq: 20 Oct 2022 16:43

Tgt Ion:128 Resp: 58197
Ion Ratio Lower Upper
128 100
127 12.7 11.0 16.4
129 10.7 8.4 12.6

