

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102023\
 Data File : VY016059.D
 Acq On : 20 Oct 2023 12:17
 Operator : SY/MD
 Sample : 04974-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-4RE

Quant Time: Oct 20 23:05:59 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
 Quant Title : SW846 8260
 QLast Update : Thu Oct 19 03:15:44 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	9043	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	14356	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	14455	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	6777	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	7097	69.050	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 138.100%	
35) Dibromofluoromethane	7.722	113	5567	58.914	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 117.820%	
50) Toluene-d8	10.179	98	17498	50.987	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.980%	
62) 4-Bromofluorobenzene	12.483	95	6219	50.787	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.580%	
Target Compounds						Qvalue
20) Methylene Chloride	4.723	84	1652	14.596	ug/l	# 78

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102023\
Data File : VY016059.D
Acq On : 20 Oct 2023 12:17
Operator : SY/MD
Sample : 04974-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-4RE

Quant Time: Oct 20 23:05:59 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y101823S.M
Quant Title : SW846 8260
QLast Update : Thu Oct 19 03:15:44 2023
Response via : Initial Calibration

