

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102121\
 Data File : VY006425.D
 Acq On : 21 Oct 2021 15:07
 Operator : SY/MD
 Sample : VY1021SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1021SBL01

Quant Time: Oct 22 03:37:12 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
 Quant Title : SW846 8260
 QLast Update : Tue Sep 28 17:44:36 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	116681	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	186862	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	171532	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	83888	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	68384	54.759	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 109.520%	
35) Dibromofluoromethane	7.722	113	50217	50.301	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 100.600%	
50) Toluene-d8	10.179	98	196109	46.936	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 93.880%	
62) 4-Bromofluorobenzene	12.483	95	66796	47.060	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 94.120%	
Target Compounds						
20) Methylene Chloride	4.722	84	13465	8.058	ug/l	Qvalue 94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102121\
Data File : VY006425.D
Acq On : 21 Oct 2021 15:07
Operator : SY/MD
Sample : VY1021SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1021SBL01

Quant Time: Oct 22 03:37:12 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y092821S.M
Quant Title : SW846 8260
QLast Update : Tue Sep 28 17:44:36 2021
Response via : Initial Calibration

