

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102221\
 Data File : VY006439.D
 Acq On : 22 Oct 2021 14:46
 Operator : SY/MD
 Sample : VY1022SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1022SBL01

Quant Time: Oct 23 07:05:14 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	104222	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	178467	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	166956	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	75923	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	76364	59.542	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.080%
35) Dibromofluoromethane	7.728	113	54199	54.487	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.980%
50) Toluene-d8	10.179	98	224980	53.353	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.700%
62) 4-Bromofluorobenzene	12.483	95	72145	50.790	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	101.580%
Target Compounds						
20) Methylene Chloride	4.723	84	11117	6.166	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102221\
Data File : VY006439.D
Acq On : 22 Oct 2021 14:46
Operator : SY/MD
Sample : VY1022SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1022SBL01

Quant Time: Oct 23 07:05:14 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

