

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102221\
 Data File : VY006447.D
 Acq On : 22 Oct 2021 18:22
 Operator : SY/MD
 Sample : M4315-10
 Misc : 4.45G/5ML/MSVOA_Y/SOIL
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 MH-DD-D

Quant Time: Oct 23 07:24:03 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	98792	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	173415	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	161248	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	71206	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	72773	59.861	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.720%
35) Dibromofluoromethane	7.722	113	52309	54.118	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.240%
50) Toluene-d8	10.185	98	219853	53.656	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.320%
62) 4-Bromofluorobenzene	12.483	95	67659	49.019	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.040%

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102221\
Data File : VY006447.D
Acq On : 22 Oct 2021 18:22
Operator : SY/MD
Sample : M4315-10
Misc : 4.45G/5ML/MSVOA_Y/SOIL
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
MH-DD-D

Quant Time: Oct 23 07:24:03 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

