

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102221\
 Data File : VY006450.D
 Acq On : 22 Oct 2021 19:32
 Operator : SY/MD
 Sample : M4324-02
 Misc : 5.01G/5ML/MSVOA_Y/SOIL
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FK-SF-02-007-A

Quant Time: Oct 23 07:25:00 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	103623	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	179655	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	165200	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	64492	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	78948	61.913	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 123.820%	
35) Dibromofluoromethane	7.722	113	38870	38.818	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 77.640%	
50) Toluene-d8	10.185	98	232631	54.803	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 109.600%	
62) 4-Bromofluorobenzene	12.483	95	71103	49.725	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 99.460%	
Target Compounds						Qvalue
16) Acetone	3.960	43	4612	12.561	ug/l	91
84) 1,2,4-Trimethylbenzene	13.130	105	4056	1.074	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102221\
Data File : VY006450.D
Acq On : 22 Oct 2021 19:32
Operator : SY/MD
Sample : M4324-02
Misc : 5.01G/5ML/MSVOA_Y/SOIL
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FK-SF-02-007-A

Quant Time: Oct 23 07:25:00 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

