


Data Path : Z:\voasrv\HPCHEM1\MSVOA\_Y\Data\VY102221\  
Data File : VY006431.D  
Acq On : 22 Oct 2021 10:47  
Operator : SY/MD  
Sample : BFB  
Misc : 5.00G/5ML/MSVOA\_Y/SOIL  
ALS Vial : 1 Sample Multiplier: 1

Instrument :  
MSVOA\_Y  
ClientSampleId :  
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA\_Y\methods\82Y102221S.M  
Title : SW846 8260  
Last Update : Sat Oct 23 06:59:04 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

| Target<br>Mass | Rel. to<br>Mass | Lower<br>Limit% | Upper<br>Limit% | Rel.<br>Abn% | Raw<br>Abn | Result<br>Pass/Fail |
|----------------|-----------------|-----------------|-----------------|--------------|------------|---------------------|
| 50             | 95              | 15              | 40              | 22.7         | 17182      | PASS                |
| 75             | 95              | 30              | 60              | 57.0         | 43155      | PASS                |
| 95             | 95              | 100             | 100             | 100.0        | 75691      | PASS                |
| 96             | 95              | 5               | 9               | 6.4          | 4852       | PASS                |
| 173            | 174             | 0.00            | 2               | 0.0          | 0          | PASS                |
| 174            | 95              | 50              | 100             | 87.5         | 66213      | PASS                |
| 175            | 174             | 5               | 9               | 8.0          | 5265       | PASS                |
| 176            | 174             | 95              | 101             | 96.2         | 63672      | PASS                |
| 177            | 176             | 5               | 9               | 6.4          | 4077       | PASS                |