

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102424\
 Data File : VY020011.D
 Acq On : 24 Oct 2024 15:56
 Operator : SY/MD
 Sample : P4508-03
 Misc : 4.27g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TP-3-VOC

Quant Time: Oct 25 01:41:42 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	196333	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	412525	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	380901	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	137434	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	143247	59.256	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 118.520%	
35) Dibromofluoromethane	7.634	113	138404	50.842	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 101.680%	
50) Toluene-d8	10.103	98	518350	51.563	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 103.120%	
62) 4-Bromofluorobenzene	12.402	95	161037	44.334	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 88.660%	
Target Compounds						
16) Acetone	3.879	43	4193	6.663	ug/l	# 70

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102424\
Data File : VY020011.D
Acq On : 24 Oct 2024 15:56
Operator : SY/MD
Sample : P4508-03
Misc : 4.27g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TP-3-VOC

Quant Time: Oct 25 01:41:42 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

