

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102621\
 Data File : VY006474.D
 Acq On : 26 Oct 2021 12:23
 Operator : SY/MD
 Sample : VY1026SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1026SBL01

Quant Time: Oct 27 05:16:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	113194	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	192931	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	177866	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	79935	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	84189	60.440	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	120.880%
35) Dibromofluoromethane	7.722	113	57086	53.086	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.180%
50) Toluene-d8	10.185	98	241085	52.886	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	105.780%
62) 4-Bromofluorobenzene	12.483	95	75438	49.127	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.260%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102621\
Data File : VY006474.D
Acq On : 26 Oct 2021 12:23
Operator : SY/MD
Sample : VY1026SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1026SBL01

Quant Time: Oct 27 05:16:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

