

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102621\
 Data File : VY006482.D
 Acq On : 26 Oct 2021 15:36
 Operator : SY/MD
 Sample : M4329-03
 Misc : 4.74G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PSC124042

Quant Time: Oct 27 05:26:21 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	127216	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	213064	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	200446	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	88863	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	93449	59.694	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	119.380%
35) Dibromofluoromethane	7.722	113	62826	52.904	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	105.800%
50) Toluene-d8	10.185	98	278269	55.275	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	110.560%
62) 4-Bromofluorobenzene	12.483	95	89910	53.018	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	106.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102621\
Data File : VY006482.D
Acq On : 26 Oct 2021 15:36
Operator : SY/MD
Sample : M4329-03
Misc : 4.74G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PSC124042

Quant Time: Oct 27 05:26:21 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

