

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102621\
 Data File : VY006483.D
 Acq On : 26 Oct 2021 16:00
 Operator : SY/MD
 Sample : M4290-13
 Misc : 4.47G/5ML/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TH-BBD

Quant Time: Oct 27 05:26:29 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	127014	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	223030	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	211515	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	89306	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	102086	65.315	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 130.620%	
35) Dibromofluoromethane	7.722	113	67669	54.435	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 108.880%	
50) Toluene-d8	10.185	98	292617	55.528	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 111.060%	
62) 4-Bromofluorobenzene	12.483	95	93839	52.863	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 105.720%	
Target Compounds						
20) Methylene Chloride	4.723	84	13434	6.077	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102621\
Data File : VY006483.D
Acq On : 26 Oct 2021 16:00
Operator : SY/MD
Sample : M4290-13
Misc : 4.47G/5ML/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TH-BBD

Quant Time: Oct 27 05:26:29 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

