

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102621\
 Data File : VY006490.D
 Acq On : 26 Oct 2021 18:45
 Operator : SY/MD
 Sample : M4336-09
 Misc : 6.30G/5ML/MSVOA_Y/SOIL
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 G4

Quant Time: Oct 27 05:27:25 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	121771	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	208929	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	191372	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	74088	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	96988	64.72	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	129.44%
35) Dibromofluoromethane	7.722	113	58923	50.60	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	101.20%
50) Toluene-d8	10.185	98	266246	53.93	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	107.86%
62) 4-Bromofluorobenzene	12.483	95	82083	49.36	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	98.72%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102621\
Data File : VY006490.D
Acq On : 26 Oct 2021 18:45
Operator : SY/MD
Sample : M4336-09
Misc : 6.30G/5ML/MSVOA_Y/SOIL
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
G4

Quant Time: Oct 27 05:27:25 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

