

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102623\
 Data File : VY016082.D
 Acq On : 26 Oct 2023 17:04
 Operator : SY/MD
 Sample : VY1026SBL01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1026SBL01

Quant Time: Oct 27 04:18:58 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 27 04:12:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	141817	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	237445	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	214969	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	100000	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.136	65	81820	58.021	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.040%
35) Dibromofluoromethane	7.716	113	77348	51.600	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.200%
50) Toluene-d8	10.179	98	276754	49.843	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.680%
62) 4-Bromofluorobenzene	12.483	95	92770	48.707	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	97.420%
Target Compounds						
						Qvalue
20) Methylene Chloride	4.710	84	18186	7.231	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102623\
Data File : VY016082.D
Acq On : 26 Oct 2023 17:04
Operator : SY/MD
Sample : VY1026SBL01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1026SBL01

Quant Time: Oct 27 04:18:58 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 27 04:12:29 2023
Response via : Initial Calibration

