

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102623\
 Data File : VY016090.D
 Acq On : 26 Oct 2023 20:16
 Operator : SY/MD
 Sample : 05050-08
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 B3

Quant Time: Oct 27 04:22:10 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 27 04:12:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	38001	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	62724	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	56217	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	24317	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	21546	57.020	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.040%
35) Dibromofluoromethane	7.716	113	21013	53.067	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	106.140%
50) Toluene-d8	10.185	98	71959	49.060	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	98.120%
62) 4-Bromofluorobenzene	12.483	95	21681	43.092	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	86.180%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102623\
Data File : VY016090.D
Acq On : 26 Oct 2023 20:16
Operator : SY/MD
Sample : 05050-08
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
B3

Quant Time: Oct 27 04:22:10 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 27 04:12:29 2023
Response via : Initial Calibration

