



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102623\
 Data File : VY016073.D
 Acq On : 26 Oct 2023 08:56
 Operator : SY/MD
 Sample : BFB
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
 Title : SW846 8260
 Last Update : Fri Oct 27 04:12:29 2023



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1741

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	18.9	23592	PASS
75	95	30	60	52.1	64923	PASS
95	95	100	100	100.0	124725	PASS
96	95	5	9	6.6	8235	PASS
173	174	0.00	2	0.9	874	PASS
174	95	50	100	81.7	101843	PASS
175	174	5	9	7.2	7346	PASS
176	174	95	101	96.4	98155	PASS
177	176	5	9	6.6	6504	PASS