



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102721\
 Data File : VY006496.D
 Acq On : 27 Oct 2021 08:08
 Operator : SY/MD
 Sample : BFB808
 Misc : 5.00G/10ML/MSVOA_Y/SOIL
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BFB808

Integration File: LSCINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\SFAMYL100521SMA.M
 Title : VOC Analysis
 Last Update : Thu Oct 28 05:46:40 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
50	95	15	40	24.0	21901	PASS
75	95	30	80	58.2	53171	PASS
95	95	100	100	100.0	91349	PASS
96	95	5	9	6.6	6021	PASS
173	174	0.00	2	0.0	0	PASS
174	95	50	120	86.7	79168	PASS
175	174	5	9	8.1	6449	PASS
176	174	95	105	96.7	76555	PASS
177	176	5	10	6.5	5001	PASS

Scan 1740 (12.428 min): VY006496.D\data.ms
BFB808

m/z	Abundance
40.00	112.0
44.10	132.0
206.90	67.0

Instrument :
MSVOA_Y
ClientSampleId :
BFB808

m/z	Abundance
36.10	1148.0
37.10	6111.0
38.10	5288.0
39.10	2175.0
40.00	195.0
41.20	55.0
43.20	115.0
44.10	832.0
45.10	1291.0
46.10	126.0
47.10	1258.0
48.10	757.0
49.10	4676.0
50.10	21568.0
51.10	6889.0
52.00	300.0
53.00	135.0
55.00	327.0
56.10	1624.0
57.10	3143.0
58.00	102.0
60.00	1003.0
61.00	5441.0
62.10	5275.0
63.10	4247.0
64.00	523.0
65.10	539.0
66.10	52.0
67.10	207.0
68.10	10674.0
69.10	10523.0
70.00	837.0
72.00	591.0
73.10	4912.0
74.00	17928.0
75.10	50512.0
76.10	4291.0
76.90	438.0
77.90	369.0
78.90	3732.0
80.00	1163.0
80.90	3740.0
82.00	887.0
83.10	71.0
86.20	117.0
87.00	3579.0
88.00	3719.0
90.90	461.0
92.00	2796.0
93.10	4029.0
94.10	11271.0
95.10	86224.0
96.00	5997.0
97.20	184.0
103.90	417.0
104.90	157.0
105.90	403.0
107.00	123.0
109.90	80.0
110.90	77.0
112.80	105.0
114.80	53.0
115.90	459.0
117.00	727.0
117.90	414.0
119.00	514.0
123.90	123.0
125.10	71.0
127.90	315.0
129.00	177.0
130.00	357.0
131.00	117.0
132.90	127.0
133.90	89.0
134.80	196.0
137.00	198.0
140.00	68.0
141.00	1210.0
142.10	126.0
143.00	1310.0
144.00	84.0
144.90	273.0
146.00	80.0
146.90	61.0
148.00	190.0
149.00	64.0
149.90	105.0
152.90	60.0
155.00	243.0
156.90	189.0
158.90	107.0
161.00	86.0

Instrument :
MSVOA_Y
ClientSampleId :
BFB808

170.50	84.0
171.20	196.0
172.00	609.0
174.00	67784.0
175.00	5431.0
176.00	65952.0
177.00	4361.0
178.00	195.0
191.10	67.0
192.70	89.0
193.10	128.0
207.10	99.0
248.90	72.0
265.10	91.0
281.00	902.0
282.10	337.0
283.00	245.0

Instrument :
MSVOA_Y
ClientSampleId :
BFB808

m/z	Abundance
36.10	1114.0
37.10	6116.0
38.10	5685.0
39.10	2547.0
40.10	258.0
43.10	105.0
44.10	842.0
45.10	1090.0
46.20	144.0
47.10	1518.0
48.10	770.0
49.10	5192.0
50.10	25200.0
51.10	7634.0
52.00	254.0
53.00	69.0
55.00	430.0
56.10	2068.0
57.00	3287.0
58.00	168.0
59.10	73.0
60.00	1344.0
61.10	6168.0
62.10	5973.0
63.00	4497.0
64.10	574.0
65.10	371.0
66.10	80.0
67.00	329.0
68.00	12848.0
69.10	12525.0
70.10	1038.0
72.10	758.0
73.00	5581.0
74.10	21104.0
75.10	61328.0
76.10	5050.0
77.10	738.0
78.10	498.0
78.90	4432.0
79.90	1264.0
81.00	4290.0
82.00	965.0
83.10	100.0
86.10	152.0
87.00	4571.0
88.00	4391.0
90.90	527.0
92.00	3558.0
93.00	5155.0
94.10	13484.0
95.10	103152.0
96.10	6541.0
97.00	227.0
103.10	90.0
103.80	487.0
104.90	116.0
105.80	521.0
106.80	157.0
109.90	161.0
110.80	145.0
113.00	113.0
115.00	170.0
116.00	378.0
117.00	616.0
117.80	423.0
119.00	823.0
123.90	143.0
125.00	57.0
128.00	356.0
128.80	275.0
130.00	448.0
131.00	222.0
133.10	113.0
133.90	93.0
135.10	158.0
136.90	294.0
140.00	93.0
141.00	1159.0
141.90	196.0
142.90	1353.0
143.80	125.0
145.10	320.0
145.90	191.0
146.80	145.0
147.90	240.0
148.80	90.0
149.80	172.0
152.90	61.0
154.90	246.0
157.10	164.0
159.00	180.0

Instrument :
MSVOA_Y
ClientSampleId :
BFB808

161.00	111.0
170.00	85.0
171.00	193.0
172.00	745.0
174.00	89232.0
175.00	7808.0
176.00	85840.0
177.00	5503.0
178.00	159.0
191.20	156.0
193.10	242.0
207.00	237.0
207.90	57.0
248.90	124.0
250.80	69.0
265.10	186.0
281.00	2091.0
282.10	510.0
283.20	418.0
284.20	82.0

Instrument :
MSVOA_Y
ClientSampleId :
BFB808

m/z	Abundance
36.10	836.0
37.10	4734.0
38.10	4510.0
39.10	1759.0
40.00	204.0
43.00	63.0
44.00	802.0
45.10	854.0
46.00	96.0
47.00	1147.0
48.10	718.0
49.10	4037.0
50.10	18936.0
51.10	5929.0
52.20	175.0
52.90	63.0
55.10	301.0
56.10	1562.0
57.10	2949.0
58.10	136.0
60.00	925.0
61.10	4723.0
62.10	4472.0
63.10	3550.0
64.10	461.0
65.00	382.0
67.00	371.0
68.10	9833.0
69.10	9595.0
70.00	923.0
72.00	483.0
73.00	4703.0
74.10	16976.0
75.10	47672.0
76.10	3806.0
77.10	381.0
77.90	346.0
79.00	3188.0
80.00	961.0
81.00	3503.0
82.00	936.0
83.10	91.0
86.10	153.0
87.00	3698.0
88.00	3210.0
91.00	337.0
92.10	2918.0
93.10	4124.0
94.10	11695.0
95.10	84672.0
96.10	5525.0
97.10	155.0
102.70	58.0
103.30	87.0
103.90	454.0
104.90	162.0
105.90	492.0
106.90	84.0
109.90	135.0
110.80	159.0
112.90	131.0
114.90	97.0
115.90	437.0
116.90	663.0
117.90	519.0
119.00	761.0
124.00	121.0
125.00	81.0
125.70	62.0
128.00	328.0
128.80	90.0
129.10	118.0
129.80	516.0
131.00	128.0
133.10	222.0
133.70	85.0
135.00	244.0
136.90	198.0
140.00	78.0
140.90	1265.0
141.90	148.0
142.90	1644.0
144.00	81.0
144.90	243.0
145.90	125.0
146.70	70.0
148.00	309.0
148.80	78.0
149.90	92.0
152.90	60.0
154.90	179.0
157.10	201.0

Instrument :
MSVOA_Y
ClientSampleId :
BFB808

160.80	138.0
170.20	52.0
170.90	175.0
172.10	616.0
174.00	80488.0
175.10	6107.0
176.00	77872.0
177.00	5139.0
178.00	142.0
179.20	54.0
191.00	218.0
191.90	60.0
192.90	328.0
194.00	100.0
205.10	56.0
207.10	241.0
207.90	53.0
248.90	111.0
249.90	59.0
250.80	74.0
251.20	74.0
264.80	216.0
265.20	215.0
265.80	71.0
267.10	74.0
281.10	3247.0
282.00	977.0
283.00	605.0
283.90	169.0

Instrument :
MSVOA_Y
ClientSampleId :
BFB808