

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102723\
 Data File : VY016102.D
 Acq On : 27 Oct 2023 13:59
 Operator : SY/MD
 Sample : 05073-01
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-02-102523

Quant Time: Oct 30 06:08:15 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 27 04:12:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	59056	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	101081	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	97340	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	47054	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	37109	63.193	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	126.380%
35) Dibromofluoromethane	7.716	113	34587	54.201	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	108.400%
50) Toluene-d8	10.179	98	121429	51.372	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.740%
62) 4-Bromofluorobenzene	12.483	95	42554	52.483	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	104.960%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102723\
Data File : VY016102.D
Acq On : 27 Oct 2023 13:59
Operator : SY/MD
Sample : 05073-01
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-02-102523

Quant Time: Oct 30 06:08:15 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 27 04:12:29 2023
Response via : Initial Calibration

