

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102821\
 Data File : VY006513.D
 Acq On : 28 Oct 2021 14:12
 Operator : SY/MD
 Sample : M4375-04
 Misc : 5.09G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-9(2.0-2.5)

Quant Time: Oct 29 05:06:10 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	10106	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	13372	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	10370	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3626	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	5225	42.015	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	84.020%
35) Dibromofluoromethane	7.728	113	1446	19.401	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	38.800%#
50) Toluene-d8	10.179	98	16185	51.226	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.460%
62) 4-Bromofluorobenzene	12.483	95	3892	36.568	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	73.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102821\
Data File : VY006513.D
Acq On : 28 Oct 2021 14:12
Operator : SY/MD
Sample : M4375-04
Misc : 5.09G/5ML/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-9(2.0-2.5)

Quant Time: Oct 29 05:06:10 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

