

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102821\
 Data File : VY006523.D
 Acq On : 28 Oct 2021 18:08
 Operator : SY/MD
 Sample : M4375-04
 Misc : 5.10G/5ML/MSVOA_Y/SOIL
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-9(2.0-2.5)

Quant Time: Oct 29 07:05:49 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	108424	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	186042	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	171281	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	72052	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	80878	60.618	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 121.240%	
35) Dibromofluoromethane	7.722	113	56871	54.845	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 109.680%	
50) Toluene-d8	10.185	98	236675	53.842	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.680%	
62) 4-Bromofluorobenzene	12.483	95	72567	49.007	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 98.020%	
Target Compounds						
					Qvalue	
20) Methylene Chloride	4.729	84	9680	4.446	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102821\
Data File : VY006523.D
Acq On : 28 Oct 2021 18:08
Operator : SY/MD
Sample : M4375-04
Misc : 5.10G/5ML/MSVOA_Y/SOIL
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-9(2.0-2.5)

Quant Time: Oct 29 07:05:49 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

