

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102821\
 Data File : VY006524.D
 Acq On : 28 Oct 2021 18:32
 Operator : SY/MD
 Sample : M4375-05
 Misc : 5.02G/5ML/MSVOA_Y/SOIL
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 SB-10(1.0-1.5)

Quant Time: Oct 29 07:05:57 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	6579	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.691	114	8390	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	6129	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	1985	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	3203	39.563	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	79.120%
35) Dibromofluoromethane	7.716	113	752	16.081	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	32.160%#
50) Toluene-d8	10.185	98	10152	51.211	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	102.420%
62) 4-Bromofluorobenzene	12.483	95	2205	33.020	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	66.040%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102821\
Data File : VY006524.D
Acq On : 28 Oct 2021 18:32
Operator : SY/MD
Sample : M4375-05
Misc : 5.02G/5ML/MSVOA_Y/SOIL
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
SB-10(1.0-1.5)

Quant Time: Oct 29 07:05:57 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

