

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102921\
 Data File : VY006528.D
 Acq On : 29 Oct 2021 10:23
 Operator : SY/MD
 Sample : VY1029SBL01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBL01

Quant Time: Oct 30 02:55:55 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	114190	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	191878	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	178795	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	76916	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	82512	58.720	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.440%
35) Dibromofluoromethane	7.722	113	57508	53.772	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	107.540%
50) Toluene-d8	10.185	98	241788	53.332	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.660%
62) 4-Bromofluorobenzene	12.483	95	76079	49.816	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	99.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102921\
Data File : VY006528.D
Acq On : 29 Oct 2021 10:23
Operator : SY/MD
Sample : VY1029SBL01
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBL01

Quant Time: Oct 30 02:55:55 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

