

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102921\
 Data File : VY006530.D
 Acq On : 29 Oct 2021 11:22
 Operator : SY/MD
 Sample : VY1029SBS01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBS01

Manual Integrations
 APPROVED

Reviewed By :Vimala Arumugam 11/01/2021
 Supervised By :Mahesh Dadoda 11/03/2021

Quant Time: Oct 30 02:56:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	147013	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	231400	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	215118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	106366	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.143	65	91038	50.323	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 100.640%			
35) Dibromofluoromethane	7.728	113	63485	49.222	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 98.440%			
50) Toluene-d8	10.185	98	277230	50.705	ug/l	0.00
Spiked Amount 50.000	Range 49 - 140		Recovery = 101.420%			
62) 4-Bromofluorobenzene	12.483	95	92815	50.394	ug/l	0.00
Spiked Amount 50.000	Range 25 - 144		Recovery = 100.780%			
Target Compounds						
					Qvalue	
2) Dichlorodifluoromethane	1.906	85	23176	17.054	ug/l	99
3) Chloromethane	2.119	50	32515	17.726	ug/l	97
4) Vinyl Chloride	2.253	62	36875	18.000	ug/l	100
5) Bromomethane	2.644	94	28117	20.602	ug/l	98
6) Chloroethane	2.796	64	22922	18.948	ug/l	97
7) Trichlorofluoromethane	3.131	101	48442	18.623	ug/l	100
8) Diethyl Ether	3.534	74	16286	19.117	ug/l	93
9) 1,1,2-Trichlorotrifluo...	3.906	101	25934	19.583	ug/l	99
10) Methyl Iodide	4.101	142	19282	17.028	ug/l	99
11) Tert butyl alcohol	4.960	59	14719	89.559	ug/l #	84
12) 1,1-Dichloroethene	3.881	96	24875	18.742	ug/l	99
13) Acrolein	3.735	56	11264	99.139	ug/l	98
14) Allyl chloride	4.491	41	55210	20.210	ug/l	96
15) Acrylonitrile	5.174	53	46692	97.328	ug/l	99
16) Acetone	3.954	43	44721	85.850	ug/l	100
17) Carbon Disulfide	4.198	76	76440	17.559	ug/l	99
18) Methyl Acetate	4.485	43	33120	19.607	ug/l	99
19) Methyl tert-butyl Ether	5.228	73	80785	18.985	ug/l	98
20) Methylene Chloride	4.722	84	37728	20.994	ug/l	98
21) trans-1,2-Dichloroethene	5.228	96	27833	18.492	ug/l	94
22) Diisopropyl ether	6.125	45	114035	20.945	ug/l	98
23) Vinyl Acetate	6.064	43	328701	99.264	ug/l	98
24) 1,1-Dichloroethane	6.021	63	57796	20.302	ug/l	100
25) 2-Butanone	6.990	43	65298	95.681	ug/l	100
26) 2,2-Dichloropropane	6.990	77	53119	19.617	ug/l	99
27) cis-1,2-Dichloroethene	6.996	96	32623	18.986	ug/l	96
28) Bromochloromethane	7.338	49	26602	23.031	ug/l	91
29) Tetrahydrofuran	7.350	42	42335	101.112	ug/l	95
30) Chloroform	7.509	83	57275	19.482	ug/l	99
31) Cyclohexane	7.789	56	55093	18.811	ug/l	96
32) 1,1,1-Trichloroethane	7.704	97	51706	19.047	ug/l	99
36) 1,1-Dichloropropene	7.923	75	42912	18.768	ug/l	98
37) Ethyl Acetate	7.076	43	28650	19.800	ug/l	100
38) Carbon Tetrachloride	7.905	117	45949	18.217	ug/l	95
39) Methylcyclohexane	9.185	83	50436	18.121	ug/l	95
40) Benzene	8.167	78	121476	18.628	ug/l	98

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102921\
 Data File : VY006530.D
 Acq On : 29 Oct 2021 11:22
 Operator : SY/MD
 Sample : VY1029SBS01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBS01

Manual Integrations
 APPROVED

Reviewed By :Vimala Arumugam 11/01/2021
 Supervised By :Mahesh Dadoda 11/03/2021

Quant Time: Oct 30 02:56:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.313	41	15712m	19.136	ug/l	
42) 1,2-Dichloroethane	8.240	62	43629	18.735	ug/l	98
43) Isopropyl Acetate	8.277	43	54228	19.318	ug/l	99
44) Trichloroethene	8.941	130	32167	19.126	ug/l	98
45) 1,2-Dichloropropane	9.222	63	32309	20.002	ug/l	96
46) Dibromomethane	9.307	93	16363	18.455	ug/l	98
47) Bromodichloromethane	9.502	83	43694	18.944	ug/l	98
48) Methyl methacrylate	9.295	41	25937	19.936	ug/l	92
49) 1,4-Dioxane	9.301	88	4203	349.167	ug/l	96
51) 4-Methyl-2-Pentanone	10.075	43	145067	98.585	ug/l	98
52) Toluene	10.246	92	76047	18.587	ug/l	100
53) t-1,3-Dichloropropene	10.465	75	46665	18.780	ug/l	99
54) cis-1,3-Dichloropropene	9.929	75	51391	19.090	ug/l	99
55) 1,1,2-Trichloroethane	10.648	97	23157	18.890	ug/l	93
56) Ethyl methacrylate	10.508	69	35600	18.797	ug/l	99
57) 1,3-Dichloropropane	10.795	76	42978	19.290	ug/l	99
58) 2-Chloroethyl Vinyl ether	9.783	63	88297	102.199	ug/l	98
59) 2-Hexanone	10.831	43	103287	96.984	ug/l	100
60) Dibromochloromethane	10.984	129	27932	18.472	ug/l	100
61) 1,2-Dibromoethane	11.093	107	22340	19.356	ug/l	96
64) Tetrachloroethene	10.721	164	35982	19.129	ug/l	97
65) Chlorobenzene	11.520	112	78775	18.816	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.593	131	29370	18.635	ug/l	98
67) Ethyl Benzene	11.593	91	149363	18.736	ug/l	99
68) m/p-Xylenes	11.703	106	111797	37.278	ug/l	98
69) o-Xylene	12.032	106	53470	18.685	ug/l	98
70) Styrene	12.044	104	91633	18.728	ug/l	98
71) Bromoform	12.209	173	18614	18.684	ug/l #	97
73) Isopropylbenzene	12.331	105	142162	19.107	ug/l	99
74) N-amyl acetate	12.142	43	48281	19.561	ug/l	100
75) 1,1,2,2-Tetrachloroethane	12.581	83	24226	18.807	ug/l	98
76) 1,2,3-Trichloropropane	12.636	75	22281m	20.328	ug/l	
77) Bromobenzene	12.611	156	35440	20.217	ug/l	97
78) n-propylbenzene	12.672	91	180483	19.628	ug/l	98
79) 2-Chlorotoluene	12.758	91	102302	20.096	ug/l	99
80) 1,3,5-Trimethylbenzene	12.812	105	126473	20.300	ug/l	97
81) trans-1,4-Dichloro-2-b...	12.380	75	10267	19.230	ug/l	98
82) 4-Chlorotoluene	12.855	91	107446	20.256	ug/l	99
83) tert-Butylbenzene	13.075	119	102798	19.201	ug/l	99
84) 1,2,4-Trimethylbenzene	13.123	105	119348	19.159	ug/l	99
85) sec-Butylbenzene	13.251	105	157100	19.543	ug/l	100
86) p-Isopropyltoluene	13.367	119	130294	18.999	ug/l	99
87) 1,3-Dichlorobenzene	13.367	146	66592	18.928	ug/l	99
88) 1,4-Dichlorobenzene	13.447	146	66796	19.309	ug/l	99
89) n-Butylbenzene	13.696	91	127549	19.771	ug/l	99
90) Hexachloroethane	13.959	117	24702	19.943	ug/l	96
91) 1,2-Dichlorobenzene	13.739	146	58522	18.964	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.355	75	5044	18.445	ug/l	98
93) 1,2,4-Trichlorobenzene	15.007	180	38391	18.845	ug/l	100
94) Hexachlorobutadiene	15.111	225	24823	19.012	ug/l	99
95) Naphthalene	15.239	128	72077	18.076	ug/l	99
96) 1,2,3-Trichlorobenzene	15.422	180	33037	18.519	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102921\
 Data File : VY006530.D
 Acq On : 29 Oct 2021 11:22
 Operator : SY/MD
 Sample : VY1029SBSD01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
ClientSampleId :
 VY1029SBSD01


Manual Integrations
APPROVED

Reviewed By :Vimala Arumugam 11/01/2021
 Supervised By :Mahesh Dadoda 11/03/2021

Quant Time: Oct 30 02:56:54 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.795 min Scan# 980
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBSD01




Tgt Ion:168 Resp: 147011
Ion Ratio Lower Upper
168 100
99 59.0 44.8 67.2

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#2
Dichlorodifluoromethane
Concen: 17.054 ug/l
RT: 1.906 min Scan# 14
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion: 85 Resp: 23176
Ion Ratio Lower Upper
85 100
87 31.9 15.8 47.3





#3
Chloromethane
Concen: 17.726 ug/l
RT: 2.119 min Scan# 49
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 50 Resp: 3251
Ion Ratio Lower Upper
50 100
52 30.2 25.4 38.2




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#4
Vinyl Chloride
Concen: 18.000 ug/l
RT: 2.253 min Scan# 71
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Tgt Ion: 62 Resp: 36875
Ion Ratio Lower Upper
62 100
64 31.2 24.7 37.1









#5
Bromomethane
Concen: 20.602 ug/l
RT: 2.644 min Scan# 11
Delta R.T. -0.012 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion: 94 Resp: 2811
Ion Ratio Lower Upper
94 100
96 89.6 73.5 110.3




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#6
Chloroethane
Concen: 18.948 ug/l
RT: 2.796 min Scan# 160
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Tgt Ion: 64 Resp: 22922
Ion Ratio Lower Upper
64 100
66 31.1 23.8 35.6









#7
Trichlorofluoromethane
Concen: 18.623 ug/l
RT: 3.131 min Scan# 216
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion:101 Resp: 4844
Ion Ratio Lower Upper
101 100
103 62.7 50.2 75.2




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#8
Diethyl Ether
Concen: 19.117 ug/l
RT: 3.534 min Scan# 281
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Tgt Ion: 74 Resp: 16286
Ion Ratio Lower Upper
74 100
45 112.6 52.8 158.3









#9
1,1,2-Trichlorotrifluoroethane
Concen: 19.583 ug/l
RT: 3.906 min Scan# 343
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion:	101	Resp:	25934
Ion	Ratio	Lower	Upper
101	100		
85	47.8	38.6	57.8
151	79.0	64.0	96.0




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#10
Methyl Iodide
Concen: 17.028 ug/l
RT: 4.101 min Scan# 374
Delta R.T. -0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion:	142	Resp:	19282
Ion	Ratio	Lower	Upper
142	100		
127	49.7	40.1	60.1
141	14.3	12.5	18.7





#11
Tert butyl alcohol
Concen: 89.559 ug/l
RT: 4.960 min Scan# 515
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion: 59 Resp: 14719
Ion Ratio Lower Upper
59 100
57 4.4 8.2 12.4

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#12
1,1-Dichloroethene
Concen: 18.742 ug/l
RT: 3.881 min Scan# 338
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion: 96 Resp: 24875
Ion Ratio Lower Upper
96 100
61 178.8 143.0 214.6
98 64.9 50.5 75.7





#13
Acrolein
Concen: 99.139 ug/l
RT: 3.735 min Scan# 314
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 56 Resp: 11264
Ion Ratio Lower Upper
56 100
55 69.9 57.1 85.7




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#14
Allyl chloride
Concen: 20.210 ug/l
RT: 4.491 min Scan# 438
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 41 Resp: 55210
Ion Ratio Lower Upper
41 100
39 69.9 58.8 88.2
76 30.0 25.0 37.4





#15
Acrylonitrile
Concen: 97.328 ug/l
RT: 5.174 min Scan# 51
Delta R.T. 0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion: 53 Resp: 46692
Ion Ratio Lower Upper
53 100
52 85.1 67.1 100.7
51 38.0 30.0 45.0

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#16
Acetone
Concen: 85.850 ug/l
RT: 3.954 min Scan# 350
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion: 43 Resp: 44721
Ion Ratio Lower Upper
43 100
58 28.4 22.9 34.3





#17
Carbon Disulfide
Concen: 17.559 ug/l
RT: 4.198 min Scan# 391
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 76 Resp: 76440
Ion Ratio Lower Upper
76 100
78 8.7 7.1 10.7




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#18
Methyl Acetate
Concen: 19.607 ug/l
RT: 4.485 min Scan# 437
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 43 Resp: 33120
Ion Ratio Lower Upper
43 100
74 20.4 16.6 25.0





#19
Methyl tert-butyl Ether
Concen: 18.985 ug/l
RT: 5.228 min Scan# 51
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 73 Resp: 8078
Ion Ratio Lower Upper
73 100
57 23.8 18.2 27.2




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#20
Methylene Chloride
Concen: 20.994 ug/l
RT: 4.722 min Scan# 476
Delta R.T. -0.007 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 84 Resp: 37728
Ion Ratio Lower Upper
84 100
49 134.5 107.8 161.8
51 42.3 32.8 49.2
86 60.1 50.6 75.8





#21
trans-1,2-Dichloroethene
Concen: 18.492 ug/l
RT: 5.228 min Scan# 511
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion: 96 Resp: 2783
Ion Ratio Lower Upper
96 100
61 143.9 120.6 181.0
98 59.7 51.3 76.9

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#22
Diisopropyl ether
Concen: 20.945 ug/l
RT: 6.125 min Scan# 706
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion: 45 Resp: 114035
Ion Ratio Lower Upper
45 100
43 61.3 47.7 71.5
87 23.6 20.1 30.1
59 10.8 8.3 12.5





#23
 Vinyl Acetate
 Concen: 99.264 ug/l
 RT: 6.064 min Scan# 61
 Delta R.T. -0.006 min
 Lab File: VY006530.D
 Acq: 29 Oct 2021 11:22




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBS01



Tgt Ion: 43 Resp: 32870
 Ion Ratio Lower Upper
 43 100
 86 8.3 7.3 10.9




Manual Integrations
 APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
 Supervised By :Mahesh Dadoda 11/03/2021



#24
 1,1-Dichloroethane
 Concen: 20.302 ug/l
 RT: 6.021 min Scan# 689
 Delta R.T. -0.006 min
 Lab File: VY006530.D
 Acq: 29 Oct 2021 11:22

Tgt Ion: 63 Resp: 57796
 Ion Ratio Lower Upper
 63 100
 98 6.7 3.5 10.4
 100 4.0 2.0 6.0









#25
2-Butanone
Concen: 95.681 ug/l
RT: 6.990 min Scan# 848
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion: 43 Resp: 65298
Ion Ratio Lower Upper
43 100
72 23.8 18.9 28.3




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#26
2,2-Dichloropropane
Concen: 19.617 ug/l
RT: 6.990 min Scan# 848
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 77 Resp: 53119
Ion Ratio Lower Upper
77 100
97 20.8 10.6 31.8





#27
cis-1,2-Dichloroethene
Concen: 18.986 ug/l
RT: 6.996 min Scan# 848
Delta R.T. 0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion: 96 Resp: 3262
Ion Ratio Lower Upper
96 100
61 164.4 0.0 314.0
98 64.6 0.0 127.8

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#28
Bromochloromethane
Concen: 23.031 ug/l
RT: 7.338 min Scan# 905
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22



Tgt Ion: 49 Resp: 26602
Ion Ratio Lower Upper
49 100
129 0.6 0.0 4.4
130 62.4 55.6 83.4









#29
Tetrahydrofuran
Concen: 101.112 ug/l
RT: 7.350 min Scan# 908
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion: 42 Resp: 42335
Ion Ratio Lower Upper
42 100
72 36.7 32.0 48.0
71 34.7 30.3 45.5




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#30
Chloroform
Concen: 19.482 ug/l
RT: 7.509 min Scan# 933
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 83 Resp: 57275
Ion Ratio Lower Upper
83 100
85 62.1 50.2 75.4





#31
Cyclohexane
Concen: 18.811 ug/l
RT: 7.789 min Scan# 91
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion: 56 Resp: 5509
Ion Ratio Lower Upper
56 100
69 38.6 26.6 40.0
84 79.0 65.0 97.4

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#32
1,1,1-Trichloroethane
Concen: 19.047 ug/l
RT: 7.704 min Scan# 965
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22



Tgt Ion: 97 Resp: 51706
Ion Ratio Lower Upper
97 100
99 63.3 51.0 76.4
61 47.9 37.4 56.2





#33

1,2-Dichloroethane-d4

Concen: 50.323 ug/l

RT: 8.143 min Scan# 1038

Delta R.T. -0.006 min

Lab File: VY006530.D

Acq: 29 Oct 2021 11:22

Instrument :

MSVOA_Y

ClientSampleId :

VY1029SBS01

Tgt Ion: 65 Resp: 91038

Ion Ratio Lower Upper

65 100





67 49.6 0.0 98.2

Manual Integrations

APPROVED

Reviewed By :Vimala Arumugam 11/01/2021

Supervised By :Mahesh Dadoda 11/03/2021



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.691 min Scan# 1127

Delta R.T. -0.006 min

Lab File: VY006530.D

Acq: 29 Oct 2021 11:22




Tgt Ion:114 Resp: 231400


Ion Ratio Lower Upper

114 100

63 23.3 0.0 45.0

88 16.2 0.0 30.6









#35
Dibromofluoromethane
Concen: 49.222 ug/l
RT: 7.728 min Scan# 968
Delta R.T. 0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion:113 Resp: 6348
Ion Ratio Lower Upper
113 100
111 104.4 82.3 123.5
192 22.2 17.4 26.2




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#36
1,1-Dichloropropene
Concen: 18.768 ug/l
RT: 7.923 min Scan# 1001
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 75 Resp: 42912
Ion Ratio Lower Upper
75 100
110 33.5 17.1 51.3
77 30.1 24.9 37.3





#37
Ethyl Acetate
Concen: 19.800 ug/l
RT: 7.076 min Scan# 862
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 43 Resp: 28650
Ion Ratio Lower Upper
43 100
61 11.8 9.4 14.2
70 8.8 6.9 10.3




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#38
Carbon Tetrachloride
Concen: 18.217 ug/l
RT: 7.905 min Scan# 998
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion:117 Resp: 45949
Ion Ratio Lower Upper
117 100
119 102.3 76.6 115.0
121 31.3 24.7 37.1





#39
Methylcyclohexane
Concen: 18.121 ug/l
RT: 9.185 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 83 Resp: 50430
Ion Ratio Lower Upper
83 100
55 86.5 65.3 97.9
98 48.7 37.0 55.4




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#40
Benzene
Concen: 18.628 ug/l
RT: 8.167 min Scan# 1041
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 78 Resp: 121476
Ion Ratio Lower Upper
78 100
77 25.5 19.7 29.5





#41
Methacrylonitrile
Concen: 19.136 ug/l m
RT: 7.313 min Scan# 901
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBSD01



Tgt Ion: 41 Resp: 15712
Ion Ratio Lower Upper
41 100
39 55.0 109.0 163.6
67 0.0 0.0 0.0
52 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#42
1,2-Dichloroethane
Concen: 18.735 ug/l
RT: 8.240 min Scan# 1053
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 62 Resp: 43629
Ion Ratio Lower Upper
62 100
98 8.2 0.0 15.2





#43
Isopropyl Acetate
Concen: 19.318 ug/l
RT: 8.277 min Scan# 1059
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 43 Resp: 54228
Ion Ratio Lower Upper
43 100
61 26.0 20.5 30.7
87 11.0 9.0 13.6




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#44
Trichloroethene
Concen: 19.126 ug/l
RT: 8.941 min Scan# 1168
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Tgt Ion:130 Resp: 32167
Ion Ratio Lower Upper
130 100
95 98.0 0.0 200.4





#45

1,2-Dichloropropane

Concen: 20.002 ug/l

RT: 9.222 min Scan# 1214

Delta R.T. 0.000 min

Lab File: VY006530.D

Acq: 29 Oct 2021 11:22

Instrument :

MSVOA_Y

ClientSampleId :

VY1029SBS01

Tgt Ion: 63 Resp: 32309

Ion Ratio Lower Upper

63 100





65 28.0 24.2 36.2

Manual Integrations

APPROVED

Reviewed By :Vimala Arumugam 11/01/2021

Supervised By :Mahesh Dadoda 11/03/2021



#46

Dibromomethane

Concen: 18.455 ug/l

RT: 9.307 min Scan# 1228

Delta R.T. 0.000 min

Lab File: VY006530.D

Acq: 29 Oct 2021 11:22




Tgt Ion: 93 Resp: 16363


Ion Ratio Lower Upper

93 100

95 82.7 65.5 98.3


174 103.1 84.3 126.5





#47
Bromodichloromethane
Concen: 18.944 ug/l
RT: 9.502 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion: 83 Resp: 43694
Ion Ratio Lower Upper
83 100
85 60.5 49.9 74.9
127 8.3 7.1 10.7

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#48
Methyl methacrylate
Concen: 19.936 ug/l
RT: 9.295 min Scan# 1226
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion: 41 Resp: 25937
Ion Ratio Lower Upper
41 100
69 69.7 60.4 90.6
39 53.8 48.2 72.2





#49
1,4-Dioxane
Concen: 349.167 ug/l
RT: 9.301 min Scan# 1227
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBSD01



Tgt Ion: 88 Resp: 420
Ion Ratio Lower Upper
88 100
43 42.9 34.1 51.1
58 76.5 57.5 86.3




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#50
Toluene-d8
Concen: 50.705 ug/l
RT: 10.185 min Scan# 1372
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Tgt Ion: 98 Resp: 277230
Ion Ratio Lower Upper
98 100
100 62.9 50.9 76.3









#51
4-Methyl-2-Pentanone
Concen: 98.585 ug/l
RT: 10.075 min Scan# 1354
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion: 43 Resp: 14506
Ion Ratio Lower Upper
43 100
58 34.7 28.7 43.1




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#52
Toluene
Concen: 18.587 ug/l
RT: 10.246 min Scan# 1382
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 92 Resp: 76047
Ion Ratio Lower Upper
92 100
91 175.1 139.7 209.5





#53
t-1,3-Dichloropropene
Concen: 18.780 ug/l
RT: 10.465 min Scan# 1418
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 75 Resp: 4666
Ion Ratio Lower Upper
75 100
77 30.8 24.9 37.3




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#54
cis-1,3-Dichloropropene
Concen: 19.090 ug/l
RT: 9.929 min Scan# 1330
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 75 Resp: 51391
Ion Ratio Lower Upper
75 100
77 31.2 25.4 38.2
39 61.6 50.2 75.2





#55
1,1,2-Trichloroethane
Concen: 18.890 ug/l
RT: 10.648 min Scan# 1448
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01






Tgt Ion: 97 Resp: 2315


Ion	Ratio	Lower	Upper
97	100		
83	89.7	66.2	99.2
85	59.0	42.2	63.4
99	63.2	48.6	73.0

Manual Integrations
APPROVED

Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021






#56
Ethyl methacrylate
Concen: 18.797 ug/l
RT: 10.508 min Scan# 1425
Delta R.T. -0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22



Tgt Ion: 69 Resp: 35600


Ion	Ratio	Lower	Upper
69	100		
41	78.6	62.9	94.3
39	42.4	35.7	53.5





#57
1,3-Dichloropropane
Concen: 19.290 ug/l
RT: 10.795 min Scan# 1472
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion: 76 Resp: 42978
Ion Ratio Lower Upper
76 100
78 31.3 25.4 38.2

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#58
2-Chloroethyl Vinyl ether
Concen: 102.199 ug/l
RT: 9.783 min Scan# 1306
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion: 63 Resp: 88297
Ion Ratio Lower Upper
63 100
106 26.0 21.4 32.2





#59
2-Hexanone
Concen: 96.984 ug/l
RT: 10.831 min Scan# 1478
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion: 43 Resp: 10328
Ion Ratio Lower Upper
43 100
58 49.1 24.6 73.6

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#60
Dibromochloromethane
Concen: 18.472 ug/l
RT: 10.984 min Scan# 1503
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion:129 Resp: 27932
Ion Ratio Lower Upper
129 100
127 78.5 39.4 118.1





#61
1,2-Dibromoethane
Concen: 19.356 ug/l
RT: 11.093 min Scan# 1521
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion:107 Resp: 22340
Ion Ratio Lower Upper
107 100
109 91.0 76.2 114.2

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#62
4-Bromofluorobenzene
Concen: 50.394 ug/l
RT: 12.483 min Scan# 1749
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion: 95 Resp: 92815
Ion Ratio Lower Upper
95 100
174 86.9 0.0 177.0
176 82.5 0.0 169.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.490 min Scan# 1586
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion:117 Resp: 215118
Ion Ratio Lower Upper
117 100
82 57.6 45.7 68.5
119 31.8 25.5 38.3

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#64
Tetrachloroethene
Concen: 19.129 ug/l
RT: 10.721 min Scan# 1460
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion:164 Resp: 35982
Ion Ratio Lower Upper
164 100
166 129.5 103.3 154.9
129 87.2 71.2 106.8
131 91.7 67.6 101.4





#65
Chlorobenzene
Concen: 18.816 ug/l
RT: 11.520 min Scan# 1591
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion:112 Resp: 78775
Ion Ratio Lower Upper
112 100
114 31.8 25.8 38.8

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#66
1,1,1,2-Tetrachloroethane
Concen: 18.635 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion:131 Resp: 29370
Ion Ratio Lower Upper
131 100
133 96.7 47.8 143.3
119 64.0 30.8 92.4





#67
Ethyl Benzene
Concen: 18.736 ug/l
RT: 11.593 min Scan# 1603
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 91 Resp: 14936
Ion Ratio Lower Upper
91 100
106 30.6 24.2 36.4




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#68
m/p-Xylenes
Concen: 37.278 ug/l
RT: 11.703 min Scan# 1621
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Tgt Ion:106 Resp: 111797
Ion Ratio Lower Upper
106 100
91 206.4 167.5 251.3









#69
o-Xylene
Concen: 18.685 ug/l
RT: 12.032 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion:106 Resp: 53470
Ion Ratio Lower Upper
106 100
91 217.2 110.3 330.9




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#70
Styrene
Concen: 18.728 ug/l
RT: 12.044 min Scan# 1677
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion:104 Resp: 91633
Ion Ratio Lower Upper
104 100
78 51.6 42.6 63.8
103 54.9 44.6 67.0





#71
Bromoform
Concen: 18.684 ug/l
RT: 12.209 min Scan# 1704
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion:173 Resp: 18614
Ion Ratio Lower Upper
173 100
175 47.2 24.8 74.3
254 0.0 0.1 0.1

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.428 min Scan# 1904
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion:152 Resp: 106366
Ion Ratio Lower Upper
152 100
115 58.5 28.5 85.5
150 165.8 0.0 348.2





#73
Isopropylbenzene
Concen: 19.107 ug/l
RT: 12.331 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBSD01




Tgt Ion:105 Resp: 142162
Ion Ratio Lower Upper
105 100
120 26.0 13.3 39.9

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#74
N-ethyl acetate
Concen: 19.561 ug/l
RT: 12.142 min Scan# 1693
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion: 43 Resp: 48281
Ion Ratio Lower Upper
43 100
70 36.8 29.4 44.0
55 22.5 18.1 27.1
61 21.6 17.6 26.4





#75
1,1,2,2-Tetrachloroethane
Concen: 18.807 ug/l
RT: 12.581 min Scan# 1765
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion: 83 Resp: 24220
Ion Ratio Lower Upper
83 100
131 11.2 5.5 16.4
85 62.5 31.9 95.9

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#76
1,2,3-Trichloropropane
Concen: 20.328 ug/l m
RT: 12.636 min Scan# 1774
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22



Tgt Ion: 75 Resp: 22281
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0









#77
Bromobenzene
Concen: 20.217 ug/l
RT: 12.611 min Scan# 1770
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion:156 Resp: 35440
Ion Ratio Lower Upper
156 100
77 182.7 94.2 282.4
158 98.7 48.6 145.8




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#78
n-propylbenzene
Concen: 19.628 ug/l
RT: 12.672 min Scan# 1780
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Tgt Ion: 91 Resp: 180483
Ion Ratio Lower Upper
91 100
120 23.2 11.3 33.8









#79
2-Chlorotoluene
Concen: 20.096 ug/l
RT: 12.758 min Scan# 1794
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion: 91 Resp: 102302
Ion Ratio Lower Upper
91 100
126 34.7 17.1 51.2




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#80
1,3,5-Trimethylbenzene
Concen: 20.300 ug/l
RT: 12.812 min Scan# 1803
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion:105 Resp: 126473
Ion Ratio Lower Upper
105 100
120 50.2 24.1 72.4





#81
trans-1,4-Dichloro-2-butene
Concen: 19.230 ug/l
RT: 12.380 min Scan# 1732
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01



Tgt Ion: 75 Resp: 1026
Ion Ratio Lower Upper
75 100
53 106.2 85.8 128.6
89 45.9 39.3 58.9




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#82
4-Chlorotoluene
Concen: 20.256 ug/l
RT: 12.855 min Scan# 1810
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 91 Resp: 107446
Ion Ratio Lower Upper
91 100
126 33.9 16.8 50.4





#83
tert-Butylbenzene
Concen: 19.201 ug/l
RT: 13.075 min Scan# 1846
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion:119 Resp: 102798
Ion Ratio Lower Upper
119 100
91 68.7 34.8 104.5
134 25.5 13.0 38.9

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#84
1,2,4-Trimethylbenzene
Concen: 19.159 ug/l
RT: 13.123 min Scan# 1854
Delta R.T. 0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion:105 Resp: 119348
Ion Ratio Lower Upper
105 100
120 44.4 22.6 67.8





#85
 sec-Butylbenzene
 Concen: 19.543 ug/l
 RT: 13.251 min Scan# 1875
 Delta R.T. 0.000 min
 Lab File: VY006530.D
 Acq: 29 Oct 2021 11:22




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBSD01



Tgt Ion:105 Resp: 157100
 Ion Ratio Lower Upper
 105 100
 134 20.2 10.1 30.3




Manual Integrations
 APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
 Supervised By :Mahesh Dadoda 11/03/2021



#86
 p-Isopropyltoluene
 Concen: 18.999 ug/l
 RT: 13.367 min Scan# 1894
 Delta R.T. 0.000 min
 Lab File: VY006530.D
 Acq: 29 Oct 2021 11:22


Tgt Ion:119 Resp: 130294
 Ion Ratio Lower Upper
 119 100
 134 26.7 13.1 39.1
 91 25.4 12.4 37.4





#87
1,3-Dichlorobenzene
Concen: 18.928 ug/l
RT: 13.367 min Scan# 1894
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion:146 Resp: 66592
Ion Ratio Lower Upper
146 100
111 38.3 19.3 57.9
148 64.5 31.9 95.7

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#88
1,4-Dichlorobenzene
Concen: 19.309 ug/l
RT: 13.447 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22



Tgt Ion:146 Resp: 66796
Ion Ratio Lower Upper
146 100
111 39.3 18.9 56.5
148 64.0 32.0 96.2









#89
n-Butylbenzene
Concen: 19.771 ug/l
RT: 13.696 min Scan# 1948
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBSD01

Tgt Ion: 91 Resp: 127549
Ion Ratio Lower Upper
91 100
92 52.0 26.3 78.9
134 25.4 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#90
Hexachloroethane
Concen: 19.943 ug/l
RT: 13.959 min Scan# 1991
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Tgt Ion:117 Resp: 24702
Ion Ratio Lower Upper
117 100
201 82.7 43.4 130.1









#91
1,2-Dichlorobenzene
Concen: 18.964 ug/l
RT: 13.739 min Scan# 1955
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01

Tgt Ion:146 Resp: 5852
Ion Ratio Lower Upper
146 100
111 40.6 20.1 60.3
148 64.5 31.9 95.7




Manual Integrations
APPROVED


Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#92
1,2-Dibromo-3-Chloropropane
Concen: 18.445 ug/l
RT: 14.355 min Scan# 2056
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22


Tgt Ion: 75 Resp: 5044
Ion Ratio Lower Upper
75 100
155 78.0 39.0 116.9
157 101.1 49.2 147.6





#93
1,2,4-Trichlorobenzene
Concen: 18.845 ug/l
RT: 15.007 min Scan# 2163
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion:180 Resp: 3839
Ion Ratio Lower Upper
180 100
182 94.7 47.1 141.4
145 29.1 14.5 43.6

Manual Integrations
APPROVED



Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021




#94
Hexachlorobutadiene
Concen: 19.012 ug/l
RT: 15.111 min Scan# 2180
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Tgt Ion:225 Resp: 24823
Ion Ratio Lower Upper
225 100
223 62.8 31.6 94.7
227 63.2 31.9 95.9





#95
Naphthalene
Concen: 18.076 ug/l
RT: 15.239 min Scan# 21
Delta R.T. 0.006 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS01




Tgt Ion:128 Resp: 7207
Ion Ratio Lower Upper
128 100
127 13.4 11.0 16.4
129 10.9 8.9 13.3

Manual Integrations
APPROVED

Reviewed By :Vimala Arumugam 11/01/2021
Supervised By :Mahesh Dadoda 11/03/2021



#96
1,2,3-Trichlorobenzene
Concen: 18.519 ug/l
RT: 15.422 min Scan# 2231
Delta R.T. 0.000 min
Lab File: VY006530.D
Acq: 29 Oct 2021 11:22



Tgt Ion:180 Resp: 33037
Ion Ratio Lower Upper
180 100
182 94.2 47.8 143.4
145 30.2 15.3 45.8

