



Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102921\
Data File : VY006526.D
Acq On : 29 Oct 2021 09:09
Operator : SY/MD
Sample : BFB
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 1 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y1022215.M
Title : SW846 8260
Last Update : Sat Oct 23 06:59:04 2021



AutoFind: Scans 1748, 1749, 1750; Background Corrected with Scan 1740

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	19.9	17100	PASS
75	95	30	60	54.0	46315	PASS
95	95	100	100	100.0	85747	PASS
96	95	5	9	6.4	5530	PASS
173	174	0.00	2	0.7	531	PASS
174	95	50	100	92.3	79133	PASS
175	174	5	9	8.3	6586	PASS
176	174	95	101	97.6	77211	PASS
177	176	5	9	6.7	5173	PASS