

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
 Data File : VY020053.D
 Acq On : 29 Oct 2024 12:05
 Operator : SY/MD
 Sample : P4566-01
 Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-01-102524

Quant Time: Oct 30 01:22:27 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	182042	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.622	114	386756	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.420	117	349622	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	121448	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	141142	62.968	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 125.940%	
35) Dibromofluoromethane	7.640	113	130877	51.281	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 102.560%	
50) Toluene-d8	10.109	98	479693	50.897	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 101.800%	
62) 4-Bromofluorobenzene	12.407	95	143201	42.050	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 84.100%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
Data File : VY020053.D
Acq On : 29 Oct 2024 12:05
Operator : SY/MD
Sample : P4566-01
Misc : 5.09g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-01-102524

Quant Time: Oct 30 01:22:27 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

