

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
 Data File : VY020054.D
 Acq On : 29 Oct 2024 12:28
 Operator : SY/MD
 Sample : P4597-01
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 RED-1-1

Quant Time: Oct 30 01:22:50 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	187400	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	392865	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	363083	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	129061	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	142540	61.774	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	123.540%
35) Dibromofluoromethane	7.640	113	133922	51.658	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	103.320%
50) Toluene-d8	10.109	98	489289	51.108	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.220%
62) 4-Bromofluorobenzene	12.408	95	149511	43.221	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	=	86.440%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
Data File : VY020054.D
Acq On : 29 Oct 2024 12:28
Operator : SY/MD
Sample : P4597-01
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL/A
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
RED-1-1

Quant Time: Oct 30 01:22:50 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

