

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
 Data File : VY020055.D
 Acq On : 29 Oct 2024 12:52
 Operator : SY/MD
 Sample : P4575-01RE
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PL-02-102424RE

Quant Time: Oct 29 13:15:43 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	93471	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.615	114	188201	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	160424	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.346	152	51132	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.061	65	65450	56.868	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 113.740%	
35) Dibromofluoromethane	7.634	113	62077	49.985	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 99.960%	
50) Toluene-d8	10.109	98	224123	48.869	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 97.740%	
62) 4-Bromofluorobenzene	12.407	95	62520	37.728	ug/l	0.00
Spiked Amount	50.000	Range	29 - 146	Recovery	= 75.460%	

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
Data File : VY020055.D
Acq On : 29 Oct 2024 12:52
Operator : SY/MD
Sample : P4575-01RE
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL/B
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PL-02-102424RE

Quant Time: Oct 29 13:15:43 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

