

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
 Data File : VY020052.D
 Acq On : 29 Oct 2024 11:27
 Operator : SY/MD
 Sample : VY1029SBS02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBS02

Quant Time: Oct 30 01:21:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.713	168	192041	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	8.616	114	336078	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.414	117	294729	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.347	152	140042	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.067	65	131406	55.572	ug/l	0.00
Spiked Amount 50.000	Range 50 - 163		Recovery = 111.140%			
35) Dibromofluoromethane	7.640	113	122522	55.246	ug/l	0.00
Spiked Amount 50.000	Range 54 - 147		Recovery = 110.500%			
50) Toluene-d8	10.109	98	439117	53.617	ug/l	0.00
Spiked Amount 50.000	Range 58 - 134		Recovery = 107.240%			
62) 4-Bromofluorobenzene	12.408	95	155941	52.697	ug/l	0.00
Spiked Amount 50.000	Range 29 - 146		Recovery = 105.400%			
Target Compounds						
						Qvalue
2) Dichlorodifluoromethane	1.867	85	27562	15.417	ug/l	97
3) Chloromethane	2.074	50	35297	15.170	ug/l	96
4) Vinyl Chloride	2.208	62	42155	16.464	ug/l	95
5) Bromomethane	2.592	94	28835	17.792	ug/l	100
6) Chloroethane	2.733	64	31396	18.252	ug/l	99
7) Trichlorofluoromethane	3.062	101	69201	18.163	ug/l	94
8) Diethyl Ether	3.458	74	21104	18.226	ug/l	74
9) 1,1,2-Trichlorotrifluo...	3.818	101	41246	18.612	ug/l	91
10) Methyl Iodide	4.007	142	32498	14.499	ug/l #	88
11) Tert butyl alcohol	4.842	59	29201	145.062	ug/l #	97
12) 1,1-Dichloroethene	3.793	96	33148	16.090	ug/l	86
13) Acrolein	3.659	56	12413	71.442	ug/l	97
14) Allyl chloride	4.391	41	65947	17.741	ug/l #	90
15) Acrylonitrile	5.061	53	60569	114.362	ug/l	96
16) Acetone	3.867	43	83672	135.932	ug/l #	78
17) Carbon Disulfide	4.104	76	52736	9.550	ug/l	96
18) Methyl Acetate	4.391	43	27354	20.629	ug/l #	86
19) Methyl tert-butyl Ether	5.122	73	122621	20.717	ug/l	96
20) Methylene Chloride	4.623	84	49722	21.400	ug/l	89
21) trans-1,2-Dichloroethene	5.116	96	36813	16.418	ug/l	86
22) Diisopropyl ether	6.025	45	179961	22.226	ug/l #	90
23) Vinyl Acetate	5.964	43	482500	103.947	ug/l #	91
24) 1,1-Dichloroethane	5.921	63	93498	20.773	ug/l	96
25) 2-Butanone	6.896	43	107166	130.068	ug/l #	78
26) 2,2-Dichloropropane	6.896	77	80717	20.137	ug/l	95
27) cis-1,2-Dichloroethene	6.896	96	54014	19.493	ug/l	80
28) Bromochloromethane	7.244	49	44110	22.266	ug/l #	67
29) Tetrahydrofuran	7.262	42	51101	108.624	ug/l #	82
30) Chloroform	7.427	83	100901	21.992	ug/l	97
31) Cyclohexane	7.707	56	62294	15.767	ug/l	86
32) 1,1,1-Trichloroethane	7.616	97	79637	19.806	ug/l #	92
36) 1,1-Dichloropropene	7.835	75	56618	17.720	ug/l	96
37) Ethyl Acetate	6.988	43	37875	22.734	ug/l	96
38) Carbon Tetrachloride	7.817	117	63985	18.681	ug/l	96
39) Methylcyclohexane	9.116	83	61197	15.201	ug/l	88
40) Benzene	8.085	78	184035	19.033	ug/l	100

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
 Data File : VY020052.D
 Acq On : 29 Oct 2024 11:27
 Operator : SY/MD
 Sample : VY1029SBS02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBS02

Quant Time: Oct 30 01:21:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
41) Methacrylonitrile	7.232	41	21177	23.486	ug/l #	86
42) 1,2-Dichloroethane	8.158	62	57820	20.814	ug/l	95
43) Isopropyl Acetate	8.201	43	73050	21.500	ug/l #	79
44) Trichloroethene	8.866	130	43565	18.602	ug/l	98
45) 1,2-Dichloropropane	9.146	63	53675	22.178	ug/l	96
46) Dibromomethane	9.231	93	26286	19.879	ug/l	92
47) Bromodichloromethane	9.427	83	77311	22.027	ug/l	95
48) Methyl methacrylate	9.225	41	32229	21.564	ug/l #	77
49) 1,4-Dioxane	9.231	88	6598	445.819	ug/l #	79
51) 4-Methyl-2-Pentanone	10.000	43	205007	118.293	ug/l	88
52) Toluene	10.170	92	116202	19.252	ug/l	100
53) t-1,3-Dichloropropene	10.396	75	62361	19.747	ug/l	94
54) cis-1,3-Dichloropropene	9.859	75	72464	19.447	ug/l #	84
55) 1,1,2-Trichloroethane	10.573	97	38740	22.237	ug/l	95
56) Ethyl methacrylate	10.439	69	51265	20.507	ug/l #	75
57) 1,3-Dichloropropane	10.719	76	67376	22.022	ug/l	98
58) 2-Chloroethyl Vinyl ether	9.713	63	132482	113.841	ug/l	91
59) 2-Hexanone	10.762	43	154808	125.068	ug/l	85
60) Dibromochloromethane	10.908	129	48506	21.692	ug/l	98
61) 1,2-Dibromoethane	11.018	107	32284	20.507	ug/l	95
64) Tetrachloroethene	10.646	164	36034	17.189	ug/l #	88
65) Chlorobenzene	11.438	112	132018	19.584	ug/l	99
66) 1,1,1,2-Tetrachloroethane	11.518	131	47477	20.816	ug/l	99
67) Ethyl Benzene	11.518	91	237031	19.356	ug/l	99
68) m/p-Xylenes	11.627	106	170643	37.732	ug/l	90
69) o-Xylene	11.950	106	82429	18.970	ug/l	89
70) Styrene	11.969	104	143237	19.538	ug/l	94
71) Bromoform	12.133	173	26338	21.571	ug/l #	97
73) Isopropylbenzene	12.255	105	231268	19.633	ug/l	98
74) N-amyl acetate	12.072	43	63725	20.426	ug/l	89
75) 1,1,2,2-Tetrachloroethane	12.505	83	49063	23.441	ug/l	98
76) 1,2,3-Trichloropropane	12.554	75	32639m	11.897	ug/l	
77) Bromobenzene	12.530	156	47152	18.863	ug/l	79
78) n-propylbenzene	12.597	91	289706	20.354	ug/l	97
79) 2-Chlorotoluene	12.682	91	163675	20.059	ug/l	93
80) 1,3,5-Trimethylbenzene	12.737	105	187286	19.688	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.304	75	13660	20.098	ug/l #	81
82) 4-Chlorotoluene	12.780	91	168366	20.179	ug/l	94
83) tert-Butylbenzene	12.999	119	171652	20.142	ug/l	95
84) 1,2,4-Trimethylbenzene	13.042	105	183345	19.440	ug/l	96
85) sec-Butylbenzene	13.176	105	258714	20.319	ug/l	96
86) p-Isopropyltoluene	13.292	119	204107	19.741	ug/l	96
87) 1,3-Dichlorobenzene	13.286	146	100144	19.839	ug/l	96
88) 1,4-Dichlorobenzene	13.365	146	97988	19.759	ug/l	96
89) n-Butylbenzene	13.615	91	203844	20.191	ug/l	97
90) Hexachloroethane	13.877	117	40656	20.194	ug/l	85
91) 1,2-Dichlorobenzene	13.657	146	91185	20.558	ug/l	99
92) 1,2-Dibromo-3-Chloropr...	14.273	75	7352	22.009	ug/l	71
93) 1,2,4-Trichlorobenzene	14.919	180	45405	18.256	ug/l	96
94) Hexachlorobutadiene	15.023	225	26094	18.872	ug/l	99
95) Naphthalene	15.145	128	84360	17.984	ug/l	99
96) 1,2,3-Trichlorobenzene	15.328	180	37939	18.046	ug/l	99

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
Data File : VY020052.D
Acq On : 29 Oct 2024 11:27
Operator : SY/MD
Sample : VY1029SBS02
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024

Quant Time: Oct 30 01:21:25 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
Quant Title : SW846 8260
QLast Update : Wed Oct 16 05:44:48 2024
Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
----------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY102924\
 Data File : VY020052.D
 Acq On : 29 Oct 2024 11:27
 Operator : SY/MD
 Sample : VY1029SBS02
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBS02

Quant Time: Oct 30 01:21:25 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y100924S.M
 Quant Title : SW846 8260
 QLast Update : Wed Oct 16 05:44:48 2024
 Response via : Initial Calibration

Manual Integrations APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 7.713 min Scan# 982
Delta R.T. 0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion:168 Resp: 19204
Ion Ratio Lower Upper
168 100
99 64.0 39.1 58.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#2
Dichlorodifluoromethane
Concen: 15.417 ug/l
RT: 1.867 min Scan# 24
Delta R.T. 0.001 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 85 Resp: 27562
Ion Ratio Lower Upper
85 100
87 30.3 16.0 48.0





#3
Chloromethane
Concen: 15.170 ug/l
RT: 2.074 min Scan# 50
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion: 50 Resp: 3529
Ion Ratio Lower Upper
50 100
52 29.7 25.6 38.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#4
Vinyl Chloride
Concen: 16.464 ug/l
RT: 2.208 min Scan# 80
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 62 Resp: 42155
Ion Ratio Lower Upper
62 100
64 30.0 26.1 39.1





#5
Bromomethane
Concen: 17.792 ug/l
RT: 2.592 min Scan# 144
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02




Tgt Ion: 94 Resp: 2883
Ion Ratio Lower Upper
94 100
96 92.9 74.1 111.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#6
Chloroethane
Concen: 18.252 ug/l
RT: 2.733 min Scan# 166
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27



Tgt Ion: 64 Resp: 31396
Ion Ratio Lower Upper
64 100
66 32.0 26.1 39.1









#7
Trichlorofluoromethane
Concen: 18.163 ug/l
RT: 3.062 min Scan# 21
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion:101 Resp: 6920
Ion Ratio Lower Upper
101 100
103 66.4 49.3 73.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#8
Diethyl Ether
Concen: 18.226 ug/l
RT: 3.458 min Scan# 285
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion: 74 Resp: 21104
Ion Ratio Lower Upper
74 100
45 105.7 41.3 124.1









#9
1,1,2-Trichlorotrifluoroethane
Concen: 18.612 ug/l
RT: 3.818 min Scan# 344
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion:101 Resp: 41240
Ion Ratio Lower Upper
101 100
85 50.3 34.5 51.7
151 70.8 61.7 92.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#10
Methyl Iodide
Concen: 14.499 ug/l
RT: 4.007 min Scan# 375
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion:142 Resp: 32498
Ion Ratio Lower Upper
142 100
127 51.4 32.8 49.2#
141 14.5 11.4 17.2









#11
Tert butyl alcohol
Concen: 145.062 ug/l
RT: 4.842 min Scan# 514
Delta R.T. -0.018 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 59 Resp: 2920
Ion Ratio Lower Upper
59 100
57 2.8 3.0 4.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#12
1,1-Dichloroethene
Concen: 16.090 ug/l
RT: 3.793 min Scan# 340
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 96 Resp: 33148
Ion Ratio Lower Upper
96 100
61 165.8 114.6 171.8
98 59.8 50.8 76.2





#13
Acrolein
Concen: 71.442 ug/l
RT: 3.659 min Scan# 317
Delta R.T. 0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion: 56 Resp: 1241
Ion Ratio Lower Upper
56 100
55 72.0 55.7 83.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#14
Allyl chloride
Concen: 17.741 ug/l
RT: 4.391 min Scan# 438
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 41 Resp: 65947
Ion Ratio Lower Upper
41 100
39 75.7 58.3 87.5
76 35.8 39.0 58.6





#15
Acrylonitrile
Concen: 114.362 ug/l
RT: 5.061 min Scan# 547
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02




Tgt Ion: 53 Resp: 60569
Ion Ratio Lower Upper
53 100
52 84.4 64.7 97.1
51 37.7 28.4 42.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#16
Acetone
Concen: 135.932 ug/l
RT: 3.867 min Scan# 352
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27



Tgt Ion: 43 Resp: 83672
Ion Ratio Lower Upper
43 100
58 26.2 31.7 47.5#









#17
Carbon Disulfide
Concen: 9.550 ug/l
RT: 4.104 min Scan# 391
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 76 Resp: 52730
Ion Ratio Lower Upper
76 100
78 7.1 6.8 10.2




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#18
Methyl Acetate
Concen: 20.629 ug/l
RT: 4.391 min Scan# 438
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 43 Resp: 27354
Ion Ratio Lower Upper
43 100
74 21.8 23.3 34.9#





#19
Methyl tert-butyl Ether
Concen: 20.717 ug/l
RT: 5.122 min Scan# 51
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion: 73 Resp: 12262
Ion Ratio Lower Upper
73 100
57 23.4 17.3 25.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#20
Methylene Chloride
Concen: 21.400 ug/l
RT: 4.623 min Scan# 476
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 84 Resp: 49722
Ion Ratio Lower Upper
84 100
49 121.2 84.4 126.6
51 36.8 26.0 39.0
86 57.5 49.5 74.3





#21
trans-1,2-Dichloroethene
Concen: 16.418 ug/l
RT: 5.116 min Scan# 511
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion: 96 Resp: 3681
Ion Ratio Lower Upper
96 100
61 144.7 97.5 146.3
98 64.9 50.5 75.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#22
Diisopropyl ether
Concen: 22.226 ug/l
RT: 6.025 min Scan# 706
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion: 45 Resp: 179961
Ion Ratio Lower Upper
45 100
43 61.2 42.2 63.2
87 26.3 25.0 37.6
59 10.4 10.5 15.7#









#23
 Vinyl Acetate
 Concen: 103.947 ug/l
 RT: 5.964 min Scan# 696
 Delta R.T. 0.007 min
 Lab File: VY020052.D
 Acq: 29 Oct 2024 11:27

Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBS02

Tgt Ion: 43 Resp: 482500
 Ion Ratio Lower Upper
 43 100
 86 9.5 10.6 16.0




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024



#24
 1,1-Dichloroethane
 Concen: 20.773 ug/l
 RT: 5.921 min Scan# 689
 Delta R.T. 0.007 min
 Lab File: VY020052.D
 Acq: 29 Oct 2024 11:27

Tgt Ion: 63 Resp: 93498
 Ion Ratio Lower Upper
 63 100
 98 6.6 4.0 12.2
 100 4.2 2.6 8.0









#25
2-Butanone
Concen: 130.068 ug/l
RT: 6.896 min Scan# 849
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 43 Resp: 107160
Ion Ratio Lower Upper
43 100
72 21.3 27.3 40.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#26
2,2-Dichloropropane
Concen: 20.137 ug/l
RT: 6.896 min Scan# 849
Delta R.T. 0.013 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 77 Resp: 80717
Ion Ratio Lower Upper
77 100
97 21.4 12.0 36.1





#27
 cis-1,2-Dichloroethene
 Concen: 19.493 ug/l
 RT: 6.896 min Scan# 848
 Delta R.T. 0.007 min
 Lab File: VY020052.D
 Acq: 29 Oct 2024 11:27




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBS02



Tgt Ion: 96 Resp: 54014
 Ion Ratio Lower Upper
 96 100
 61 161.0 0.0 256.2
 98 62.9 0.0 130.0




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024



#28
 Bromochloromethane
 Concen: 22.266 ug/l
 RT: 7.244 min Scan# 906
 Delta R.T. 0.000 min
 Lab File: VY020052.D
 Acq: 29 Oct 2024 11:27

Tgt Ion: 49 Resp: 44110
 Ion Ratio Lower Upper
 49 100
 129 1.7 0.0 4.6
 130 59.2 72.7 109.1#









#29
Tetrahydrofuran
Concen: 108.624 ug/l
RT: 7.262 min Scan# 908
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 42 Resp: 5110
Ion Ratio Lower Upper
42 100
72 39.9 43.0 64.4
71 38.8 40.8 61.2




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#30
Chloroform
Concen: 21.992 ug/l
RT: 7.427 min Scan# 936
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion: 83 Resp: 100901
Ion Ratio Lower Upper
83 100
85 63.0 52.2 78.4





#31

Cyclohexane

Concen: 15.767 ug/l

RT: 7.707 min Scan# 981

Delta R.T. 0.007 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Instrument :

MSVOA_Y

ClientSampleId :

VY1029SBS02

Tgt Ion: 56 Resp: 62294

Ion Ratio Lower Upper

56 100

69 38.9 28.7 43.1





84 79.7 77.7 116.5

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#32

1,1,1-Trichloroethane

Concen: 19.806 ug/l

RT: 7.616 min Scan# 967

Delta R.T. 0.000 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27




Tgt Ion: 97 Resp: 79637


Ion Ratio Lower Upper

97 100

99 65.3 50.3 75.5

61 50.6 32.6 48.8#





#33

1,2-Dichloroethane-d4

Concen: 55.572 ug/l

RT: 8.067 min Scan# 1040

Delta R.T. 0.007 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Instrument :

MSVOA_Y

ClientSampleId :

VY1029SBS02

Tgt Ion: 65 Resp: 131400

Ion Ratio Lower Upper

65 100





67 51.9 0.0 109.6

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.616 min Scan# 1131

Delta R.T. 0.000 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27




Tgt Ion:114 Resp: 336078


Ion Ratio Lower Upper

114 100

63 23.7 0.0 35.0


88 16.6 0.0 27.2





#35
Dibromofluoromethane
Concen: 55.246 ug/l
RT: 7.640 min Scan# 91
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02




Tgt Ion:113 Resp: 12252
Ion Ratio Lower Upper
113 100
111 105.6 82.2 123.4
192 16.6 15.9 23.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#36
1,1-Dichloropropene
Concen: 17.720 ug/l
RT: 7.835 min Scan# 1003
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27



Tgt Ion: 75 Resp: 56618
Ion Ratio Lower Upper
75 100
110 33.7 19.0 57.0
77 32.2 25.3 37.9









#37
Ethyl Acetate
Concen: 22.734 ug/l
RT: 6.988 min Scan# 80
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 43 Resp: 37875
Ion Ratio Lower Upper
43 100
61 13.3 12.6 19.0
70 9.6 7.8 11.8




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#38
Carbon Tetrachloride
Concen: 18.681 ug/l
RT: 7.817 min Scan# 1000
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion:117 Resp: 63985
Ion Ratio Lower Upper
117 100
119 91.3 75.8 113.6
121 32.3 24.0 36.0





#39
Methylcyclohexane
Concen: 15.201 ug/l
RT: 9.116 min Scan# 11
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion: 83 Resp: 6119
Ion Ratio Lower Upper
83 100
55 75.1 51.5 77.3
98 43.8 40.5 60.7




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#40
Benzene
Concen: 19.033 ug/l
RT: 8.085 min Scan# 1044
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion: 78 Resp: 184035
Ion Ratio Lower Upper
78 100
77 24.5 19.5 29.3









#41
Methacrylonitrile
Concen: 23.486 ug/l
RT: 7.232 min Scan# 904
Delta R.T. 0.013 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 41 Resp: 2117
Ion Ratio Lower Upper
41 100
39 62.1 48.0 72.0
67 69.7 75.1 112.7
52 34.7 29.8 44.6




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#42
1,2-Dichloroethane
Concen: 20.814 ug/l
RT: 8.158 min Scan# 1056
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 62 Resp: 57820
Ion Ratio Lower Upper
62 100
98 9.2 0.0 22.0





#43
Isopropyl Acetate
Concen: 21.500 ug/l
RT: 8.201 min Scan# 1063
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion: 43 Resp: 73050
Ion Ratio Lower Upper
43 100
61 20.0 27.9 41.9
87 12.2 13.1 19.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#44
Trichloroethene
Concen: 18.602 ug/l
RT: 8.866 min Scan# 1172
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion:130 Resp: 43565
Ion Ratio Lower Upper
130 100
95 96.4 0.0 188.0





#45

1,2-Dichloropropane

Concen: 22.178 ug/l

RT: 9.146 min Scan# 1217

Delta R.T. 0.007 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Instrument :

MSVOA_Y

ClientSampleId :

VY1029SBS02

Tgt Ion: 63 Resp: 5367

Ion Ratio Lower Upper

63 100





65 29.4 25.4 38.2

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#46

Dibromomethane

Concen: 19.879 ug/l

RT: 9.231 min Scan# 1232

Delta R.T. 0.000 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27




Tgt Ion: 93 Resp: 26286


Ion Ratio Lower Upper

93 100

95 83.5 66.2 99.4

174 87.0 80.9 121.3









#47
Bromodichloromethane
Concen: 22.027 ug/l
RT: 9.427 min Scan# 1263
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 83 Resp: 7731
Ion Ratio Lower Upper
83 100
85 66.9 50.1 75.1
127 8.6 7.4 11.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#48
Methyl methacrylate
Concen: 21.564 ug/l
RT: 9.225 min Scan# 1231
Delta R.T. 0.007 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 41 Resp: 32229
Ion Ratio Lower Upper
41 100
69 79.0 84.1 126.1#
39 67.0 43.6 65.4#





#49
1,4-Dioxane
Concen: 445.819 ug/l
RT: 9.231 min Scan# 1231
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02




Tgt Ion: 88 Resp: 6598
Ion Ratio Lower Upper
88 100
43 41.8 0.0 0.0
58 77.8 49.3 73.9

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#50
Toluene-d8
Concen: 53.617 ug/l
RT: 10.109 min Scan# 1376
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27



Tgt Ion: 98 Resp: 439117
Ion Ratio Lower Upper
98 100
100 64.3 52.0 78.0









#51
4-Methyl-2-Pentanone
Concen: 118.293 ug/l
RT: 10.000 min Scan# 1358
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 43 Resp: 20500
Ion Ratio Lower Upper
43 100
58 36.7 35.7 53.5




Manual Integrations
APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#52
Toluene
Concen: 19.252 ug/l
RT: 10.170 min Scan# 1386
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion: 92 Resp: 116202
Ion Ratio Lower Upper
92 100
91 173.4 139.0 208.6









#53
t-1,3-Dichloropropene
Concen: 19.747 ug/l
RT: 10.396 min Scan# 1422
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 75 Resp: 6236
Ion Ratio Lower Upper
75 100
77 33.6 24.1 36.1




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#54
cis-1,3-Dichloropropene
Concen: 19.447 ug/l
RT: 9.859 min Scan# 1335
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 75 Resp: 72464
Ion Ratio Lower Upper
75 100
77 32.0 25.4 38.0
39 54.2 29.8 44.6#





#55
1,1,2-Trichloroethane
Concen: 22.237 ug/l
RT: 10.573 min Scan# 1452
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02




Tgt Ion: 97	Resp: 38740
Ion Ratio	Lower Upper
97 100	
83 87.2	67.5 101.3
85 59.6	42.9 64.3
99 61.9	46.6 70.0

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#56
Ethyl methacrylate
Concen: 20.507 ug/l
RT: 10.439 min Scan# 1430
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Tgt Ion: 69	Resp: 51265
Ion Ratio	Lower Upper
69 100	
41 70.4	42.3 63.5#
39 42.0	22.9 34.3#





#57
1,3-Dichloropropane
Concen: 22.022 ug/l
RT: 10.719 min Scan# 1475
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion: 76 Resp: 67370
Ion Ratio Lower Upper
76 100
78 31.9 26.2 39.4




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#58
2-Chloroethyl Vinyl ether
Concen: 113.841 ug/l
RT: 9.713 min Scan# 1311
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 63 Resp: 132482
Ion Ratio Lower Upper
63 100
106 24.8 23.7 35.5





#59
2-Hexanone
Concen: 125.068 ug/l
RT: 10.762 min Scan# 1483
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27



Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion: 43 Resp: 154808
Ion Ratio Lower Upper
43 100
58 50.9 31.4 94.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#60
Dibromochloromethane
Concen: 21.692 ug/l
RT: 10.908 min Scan# 1507
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion:129 Resp: 48506
Ion Ratio Lower Upper
129 100
127 76.2 39.0 117.0









#61
1,2-Dibromoethane
Concen: 20.507 ug/l
RT: 11.018 min Scan# 11
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion:107 Resp: 32284
Ion Ratio Lower Upper
107 100
109 99.5 75.9 113.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#62
4-Bromofluorobenzene
Concen: 52.697 ug/l
RT: 12.408 min Scan# 1753
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion: 95 Resp: 155941
Ion Ratio Lower Upper
95 100
174 71.9 0.0 175.6
176 70.0 0.0 171.4





#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.414 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion:117 Resp: 294729
Ion Ratio Lower Upper
117 100
82 59.6 42.4 63.6
119 32.4 25.9 38.9




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#64
Tetrachloroethene
Concen: 17.189 ug/l
RT: 10.646 min Scan# 1464
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion:164 Resp: 36034
Ion Ratio Lower Upper
164 100
166 125.4 102.6 154.0
129 108.3 72.0 108.0#
131 104.6 70.0 105.0





#65
Chlorobenzene
Concen: 19.584 ug/l
RT: 11.438 min Scan# 11
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02




Tgt Ion:112 Resp: 132018
Ion Ratio Lower Upper
112 100
114 32.6 26.4 39.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#66
1,1,1,2-Tetrachloroethane
Concen: 20.816 ug/l
RT: 11.518 min Scan# 1607
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27



Tgt Ion:131 Resp: 47477
Ion Ratio Lower Upper
131 100
133 93.6 47.8 143.3
119 67.6 34.1 102.2





#67

Ethyl Benzene

Concen: 19.356 ug/l

RT: 11.518 min Scan# 1607

Delta R.T. -0.006 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Instrument :

MSVOA_Y

ClientSampleId :

VY1029SBS02

Tgt Ion: 91 Resp: 23703

Ion Ratio Lower Upper

91 100





106 30.3 24.5 36.7

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#68

m/p-Xylenes

Concen: 37.732 ug/l

RT: 11.627 min Scan# 1625

Delta R.T. -0.006 min

Lab File: VY020052.D




Acq: 29 Oct 2024 11:27


Tgt Ion:106 Resp: 170643

Ion Ratio Lower Upper

106 100

91 210.0 156.0 234.0









#69
o-Xylene
Concen: 18.970 ug/l
RT: 11.950 min Scan# 106
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion:106 Resp: 82429
Ion Ratio Lower Upper
106 100
91 223.7 103.6 310.8




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#70
Styrene
Concen: 19.538 ug/l
RT: 11.969 min Scan# 1681
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion:104 Resp: 143237
Ion Ratio Lower Upper
104 100
78 54.5 37.0 55.4
103 55.2 44.2 66.2









#71
Bromoform
Concen: 21.571 ug/l
RT: 12.133 min Scan# 1707
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion:173 Resp: 26338
Ion Ratio Lower Upper
173 100
175 48.4 25.1 75.3
254 0.0 0.0 0.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.347 min Scan# 1907
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion:152 Resp: 140042
Ion Ratio Lower Upper
152 100
115 61.6 28.2 84.7
150 164.1 0.0 345.6





#73

Isopropylbenzene

Concen: 19.633 ug/l

RT: 12.255 min Scan# 1728

Delta R.T. 0.000 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Instrument :

MSVOA_Y

ClientSampleId :

VY1029SBS02

Tgt Ion:105 Resp: 231268

Ion Ratio Lower Upper

105 100





120 25.6 13.4 40.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024

Supervised By :Mahesh Dadoda 10/30/2024



#74

N-ethyl acetate

Concen: 20.426 ug/l

RT: 12.072 min Scan# 1698

Delta R.T. 0.000 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Tgt Ion: 43 Resp: 63725




Ion Ratio Lower Upper


43 100

70 41.5 41.4 62.2

55 25.8 23.3 34.9

61 23.0 22.6 33.8









#75
1,1,2,2-Tetrachloroethane
Concen: 23.441 ug/l
RT: 12.505 min Scan# 1769
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 83 Resp: 4906
Ion Ratio Lower Upper
83 100
131 10.1 5.0 14.9
85 62.3 32.0 96.0




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#76
1,2,3-Trichloropropane
Concen: 11.897 ug/l m
RT: 12.554 min Scan# 1777
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Tgt Ion: 75 Resp: 32639
Ion Ratio Lower Upper
75 100
77 0.0 0.0 0.0





#77

Bromobenzene

Concen: 18.863 ug/l

RT: 12.530 min Scan# 1773

Delta R.T. -0.006 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Instrument :

MSVOA_Y

ClientSampleId :





VY1029SBS02

Tgt Ion:	156	Resp:	47152
Ion Ratio	Lower	Upper	
156	100		
77	216.2	87.1	261.3
158	100.8	48.6	145.9

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#78

n-propylbenzene

Concen: 20.354 ug/l




RT: 12.597 min Scan# 1784


Delta R.T. 0.000 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Tgt Ion:	91	Resp:	289706
Ion Ratio	Lower	Upper	
91	100		
120	21.6	11.5	34.4









#79
2-Chlorotoluene
Concen: 20.059 ug/l
RT: 12.682 min Scan# 1797
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27

Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02

Tgt Ion: 91 Resp: 163679
Ion Ratio Lower Upper
91 100
126 31.4 17.8 53.5




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#80
1,3,5-Trimethylbenzene
Concen: 19.688 ug/l
RT: 12.737 min Scan# 1807
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion:105 Resp: 187286
Ion Ratio Lower Upper
105 100
120 47.5 25.1 75.2





#81
trans-1,4-Dichloro-2-butene
Concen: 20.098 ug/l
RT: 12.304 min Scan# 1736
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02




Tgt Ion: 75 Resp: 13660
Ion Ratio Lower Upper
75 100
53 108.5 66.8 100.2
89 45.8 35.8 53.6

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#82
4-Chlorotoluene
Concen: 20.179 ug/l
RT: 12.780 min Scan# 1814
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Tgt Ion: 91 Resp: 168366
Ion Ratio Lower Upper
91 100
126 32.1 17.6 52.9





#83
tert-Butylbenzene
Concen: 20.142 ug/l
RT: 12.999 min Scan# 1850
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02




Tgt Ion:119 Resp: 171652
Ion Ratio Lower Upper
119 100
91 68.1 31.3 93.8
134 26.7 13.4 40.1

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#84
1,2,4-Trimethylbenzene
Concen: 19.440 ug/l
RT: 13.042 min Scan# 1857
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Tgt Ion:105 Resp: 183345
Ion Ratio Lower Upper
105 100
120 43.9 23.2 69.6





#85
 sec-Butylbenzene
 Concen: 20.319 ug/l
 RT: 13.176 min Scan# 1878
 Delta R.T. 0.000 min
 Lab File: VY020052.D
 Acq: 29 Oct 2024 11:27




Instrument :
 MSVOA_Y
 ClientSampleId :
 VY1029SBS02




Tgt Ion:105 Resp: 258714
 Ion Ratio Lower Upper
 105 100
 134 18.4 10.1 30.1

Manual Integrations
 APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024




#86
 p-Isopropyltoluene
 Concen: 19.741 ug/l
 RT: 13.292 min Scan# 1898
 Delta R.T. 0.000 min
 Lab File: VY020052.D
 Acq: 29 Oct 2024 11:27



Tgt Ion:119 Resp: 204107
 Ion Ratio Lower Upper
 119 100
 134 25.8 13.1 39.3
 91 25.2 10.7 32.1





#87

1,3-Dichlorobenzene

Concen: 19.839 ug/l

RT: 13.286 min Scan# 1897

Delta R.T. -0.006 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Instrument :

MSVOA_Y

ClientSampleId :

VY1029SBS02





Tgt Ion:146 Resp: 100144

Ion	Ratio	Lower	Upper
146	100		
111	43.0	19.0	57.0
148	62.5	31.7	95.1

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#88

1,4-Dichlorobenzene

Concen: 19.759 ug/l

RT: 13.365 min Scan# 1910




Delta R.T. -0.006 min


Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Tgt Ion:146 Resp: 97988


Ion	Ratio	Lower	Upper
146	100		
111	43.4	18.4	55.4
148	64.9	32.4	97.2





#89
n-Butylbenzene
Concen: 20.191 ug/l
RT: 13.615 min Scan# 1951
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02



Tgt Ion: 91 Resp: 20384
Ion Ratio Lower Upper
91 100
92 51.9 26.6 79.7
134 23.0 12.9 38.7




Manual Integrations
APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#90
Hexachloroethane
Concen: 20.194 ug/l
RT: 13.877 min Scan# 1994
Delta R.T. -0.006 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27


Tgt Ion:117 Resp: 40656
Ion Ratio Lower Upper
117 100
201 61.9 37.4 112.2





#91
1,2-Dichlorobenzene
Concen: 20.558 ug/l
RT: 13.657 min Scan# 1958
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27




Instrument :
MSVOA_Y
ClientSampleId :
VY1029SBS02




Tgt Ion:146 Resp: 9118
Ion Ratio Lower Upper
146 100
111 41.9 19.9 59.7
148 63.4 31.8 95.4

Manual Integrations
APPROVED



Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024




#92
1,2-Dibromo-3-Chloropropane
Concen: 22.009 ug/l
RT: 14.273 min Scan# 2059
Delta R.T. 0.000 min
Lab File: VY020052.D
Acq: 29 Oct 2024 11:27



Tgt Ion: 75 Resp: 7352
Ion Ratio Lower Upper
75 100
155 72.0 48.6 145.8
157 89.3 62.7 188.2









#93
 1,2,4-Trichlorobenzene
 Concen: 18.256 ug/l
 RT: 14.919 min Scan# 2165
 Delta R.T. 0.000 min
 Lab File: VY020052.D
 Acq: 29 Oct 2024 11:27

Instrument : MSVOA_Y
 ClientSampleId : VY1029SBS02

Tgt Ion:180	Resp: 4540
Ion Ratio	Lower Upper
180	100
182	93.9 47.9 143.8
145	34.2 14.2 42.6




Manual Integrations
 APPROVED


Reviewed By :Semsettin Yesilyurt 10/30/2024
 Supervised By :Mahesh Dadoda 10/30/2024



#94
 Hexachlorobutadiene
 Concen: 18.872 ug/l
 RT: 15.023 min Scan# 2182
 Delta R.T. 0.000 min
 Lab File: VY020052.D
 Acq: 29 Oct 2024 11:27

Tgt Ion:225	Resp: 26094
Ion Ratio	Lower Upper
225	100
223	59.5 30.4 91.2
227	63.8 31.9 95.9





#95

Naphthalene

Concen: 17.984 ug/l

RT: 15.145 min Scan# 2201

Delta R.T. 0.000 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Instrument :

MSVOA_Y

ClientSampleId :





VY1029SBS02

Tgt Ion:	128	Resp:	84360
Ion Ratio	Lower	Upper	
128	100		
127	13.4	10.6	16.0
129	11.5	8.6	13.0

Manual Integrations

APPROVED

Reviewed By :Semsettin Yesilyurt 10/30/2024
Supervised By :Mahesh Dadoda 10/30/2024



#96

1,2,3-Trichlorobenzene

Concen: 18.046 ug/l

RT: 15.328 min Scan# 2232

Delta R.T. 0.000 min

Lab File: VY020052.D

Acq: 29 Oct 2024 11:27

Tgt Ion:	180	Resp:	37939
Ion Ratio	Lower	Upper	
180	100		
182	95.4	47.6	142.9
145	32.0	15.4	46.1

