

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
 Data File : VY016126.D
 Acq On : 30 Oct 2023 12:16
 Operator : SY/MD
 Sample : 05062-01RE
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NB-07-102523RE

Quant Time: Oct 31 01:47:33 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 27 04:12:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	20907	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	33265	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	26419	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	9921	50.000	ug/l	0.01
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	9687	46.596	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	93.200%
35) Dibromofluoromethane	7.728	113	10235	48.738	ug/l	0.01
Spiked Amount	50.000	Range	54 - 147	Recovery	=	97.480%
50) Toluene-d8	10.185	98	37021	47.592	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	95.180%
62) 4-Bromofluorobenzene	12.489	95	10174	38.129	ug/l	0.01
Spiked Amount	50.000	Range	30 - 143	Recovery	=	76.260%
Target Compounds						
20) Methylene Chloride	4.710	84	2433	5.951	ug/l	84
95) Naphthalene	15.245	128	3618	10.293	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
Data File : VY016126.D
Acq On : 30 Oct 2023 12:16
Operator : SY/MD
Sample : 05062-01RE
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NB-07-102523RE

Quant Time: Oct 31 01:47:33 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 27 04:12:29 2023
Response via : Initial Calibration

