

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
 Data File : VY016127.D
 Acq On : 30 Oct 2023 12:39
 Operator : SY/MD
 Sample : 05073-01
 Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-02-102523

Quant Time: Oct 31 01:28:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 27 04:12:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	151081	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	252525	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	228433	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	110698	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	87391	58.172	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.340%
35) Dibromofluoromethane	7.716	113	81746	51.278	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	102.560%
50) Toluene-d8	10.179	98	292940	49.608	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.220%
62) 4-Bromofluorobenzene	12.483	95	100879	49.802	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.600%
Target Compounds						
20) Methylene Chloride	4.716	84	18871	6.871	ug/l	89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
Data File : VY016127.D
Acq On : 30 Oct 2023 12:39
Operator : SY/MD
Sample : 05073-01
Misc : 5.08g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-02-102523

Quant Time: Oct 31 01:28:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 27 04:12:29 2023
Response via : Initial Calibration

