

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
 Data File : VY016129.D
 Acq On : 30 Oct 2023 13:26
 Operator : SY/MD
 Sample : 05048-01
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OILY-STONE

Quant Time: Oct 31 01:29:24 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 27 04:12:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	146605	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	242941	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	219139	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	100032	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	83394	57.206	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	114.420%
35) Dibromofluoromethane	7.716	113	77386	50.458	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.920%
50) Toluene-d8	10.179	98	281789	49.602	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.200%
62) 4-Bromofluorobenzene	12.483	95	94073	48.274	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	96.540%
Target Compounds						
16) Acetone	3.948	43	9622	24.352	ug/l	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
Data File : VY016129.D
Acq On : 30 Oct 2023 13:26
Operator : SY/MD
Sample : 05048-01
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OILY-STONE

Quant Time: Oct 31 01:29:24 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 27 04:12:29 2023
Response via : Initial Calibration

