

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
 Data File : VY016130.D
 Acq On : 30 Oct 2023 13:50
 Operator : SY/MD
 Sample : 05135-02
 Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 12 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 CHRT-20413

Quant Time: Oct 31 01:29:44 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 27 04:12:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	147071	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	247928	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	227568	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	109886	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	85974	58.789	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	117.580%
35) Dibromofluoromethane	7.716	113	78678	50.268	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	100.540%
50) Toluene-d8	10.179	98	287194	49.537	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	99.080%
62) 4-Bromofluorobenzene	12.483	95	100527	50.548	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.100%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
Data File : VY016130.D
Acq On : 30 Oct 2023 13:50
Operator : SY/MD
Sample : 05135-02
Misc : 4.99g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 12 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
CHRT-20413

Quant Time: Oct 31 01:29:44 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 27 04:12:29 2023
Response via : Initial Calibration

