

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
 Data File : VY016131.D
 Acq On : 30 Oct 2023 14:14
 Operator : SY/MD
 Sample : 05145-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 TR-06-10272023

Quant Time: Oct 31 01:30:02 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 27 04:12:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	146019	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	247216	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	225916	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	109839	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	84858	58.444	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.880%
35) Dibromofluoromethane	7.716	113	58448	37.451	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	74.900%
50) Toluene-d8	10.179	98	289380	50.057	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	100.120%
62) 4-Bromofluorobenzene	12.483	95	100508	50.684	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	101.360%
Target Compounds						
						Qvalue
16) Acetone	3.948	43	10568	26.853	ug/l	93
20) Methylene Chloride	4.710	84	20004	8.176	ug/l	87

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103023\
Data File : VY016131.D
Acq On : 30 Oct 2023 14:14
Operator : SY/MD
Sample : 05145-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
TR-06-10272023

Quant Time: Oct 31 01:30:02 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102623S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 27 04:12:29 2023
Response via : Initial Calibration

