

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103122\
 Data File : VY011198.D
 Acq On : 31 Oct 2022 15:41
 Operator : KP/MD
 Sample : N5331-02RE
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ROLLOFF-RB21502

Quant Time: Nov 01 02:44:31 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	23191	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	36311	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	30744	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	10288	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	16546	68.999	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	138.000%
35) Dibromofluoromethane	7.716	113	15168	67.927	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	135.860%
50) Toluene-d8	10.179	98	47171	53.322	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.640%
62) 4-Bromofluorobenzene	12.483	95	12710	43.909	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	87.820%
Target Compounds						
20) Methylene Chloride	4.716	84	5976	20.722	ug/l	Qvalue # 82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103122\
Data File : VY011198.D
Acq On : 31 Oct 2022 15:41
Operator : KP/MD
Sample : N5331-02RE
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ROLLOFF-RB21502

Quant Time: Nov 01 02:44:31 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

