

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103122\
 Data File : VY011201.D
 Acq On : 31 Oct 2022 16:49
 Operator : KP/MD
 Sample : N5262-02RE
 Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 VOCRE

Quant Time: Nov 01 02:45:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	39431	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	63180	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	54242	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	19113	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	25969	63.693	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	127.380%	
35) Dibromofluoromethane	7.716	113	25859	66.556	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	133.120%	
50) Toluene-d8	10.179	98	79497	51.647	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	103.300%	
62) 4-Bromofluorobenzene	12.483	95	21844	43.371	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	86.740%	
Target Compounds						
20) Methylene Chloride	4.710	84	13335	29.635	ug/l	Qvalue # 79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103122\
Data File : VY011201.D
Acq On : 31 Oct 2022 16:49
Operator : KP/MD
Sample : N5262-02RE
Misc : 5.06g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 9 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
VOCRE

Quant Time: Nov 01 02:45:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

