

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103122\
 Data File : VY011207.D
 Acq On : 31 Oct 2022 19:04
 Operator : KP/MD
 Sample : N5361-01
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OR-2-102822

Quant Time: Nov 01 02:46:21 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.789	168	247759	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.685	114	431521	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.490	117	377597	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	167714	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.137	65	156990	61.279	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 122.560%	
35) Dibromofluoromethane	7.716	113	159345	60.047	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 120.100%	
50) Toluene-d8	10.179	98	564103	53.657	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	= 107.320%	
62) 4-Bromofluorobenzene	12.477	95	186441	54.199	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	= 108.400%	
Target Compounds						Qvalue
20) Methylene Chloride	4.710	84	45371	12.467	ug/l	# 84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103122\
Data File : VY011207.D
Acq On : 31 Oct 2022 19:04
Operator : KP/MD
Sample : N5361-01
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
OR-2-102822

Quant Time: Nov 01 02:46:21 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
Quant Title : SW846 8260
QLast Update : Fri Oct 28 23:42:14 2022
Response via : Initial Calibration

