

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103122\
 Data File : VY011208.D
 Acq On : 31 Oct 2022 19:27
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleID :
 VSTDCCC050

Quant Time: Nov 01 02:46:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	78	0.00
2 T	Dichlorodifluoromethane	0.272	0.236	13.2	73	0.00
3 P	Chloromethane	0.367	0.298	18.8	71	0.00
4 C	Vinyl Chloride	0.434	0.353	18.7#	70	0.00
5 T	Bromomethane	0.305	0.246	19.3	74	0.00
6 T	Chloroethane	0.300	0.266	11.3	75	0.00
7 T	Trichlorofluoromethane	0.707	0.654	7.5	78	0.00
8 T	Diethyl Ether	0.255	0.271	-6.3	91	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.446	0.435	2.5	79	0.00
10 T	Methyl Iodide	0.489	0.453	7.4	72	0.00
11 T	Tert butyl alcohol	0.057	0.067	-17.5	136	0.00
12 CM	1,1-Dichloroethene	0.402	0.372	7.5#	76	0.00
13 T	Acrolein	0.036	0.038	-5.6	107	0.00
14 T	Allyl chloride	0.624	0.649	-4.0	83	0.00
15 T	Acrylonitrile	0.140	0.168	-20.0	104	0.00
16 T	Acetone	0.126	0.126	0.0	89	0.00
17 T	Carbon Disulfide	1.043	0.713	31.6#	63	0.00
18 T	Methyl Acetate	0.400	0.390	2.5	105	0.00
19 T	Methyl tert-butyl Ether	1.186	1.383	-16.6	95	0.00
20 T	Methylene Chloride	0.656	0.627	4.4	94	0.00
21 T	trans-1,2-Dichloroethene	0.461	0.429	6.9	78	0.00
22 T	Diisopropyl ether	1.434	1.680	-17.2	91	0.00
23 T	Vinyl Acetate	0.815	1.017	-24.8	95	0.00
24 P	1,1-Dichloroethane	0.886	0.931	-5.1	85	0.00
25 T	2-Butanone	0.181	0.211	-16.6	103	0.00
26 T	2,2-Dichloropropane	0.792	0.816	-3.0	84	0.00
27 T	cis-1,2-Dichloroethene	0.558	0.580	-3.9	85	0.00
28 T	Bromochloromethane	0.405	0.460	-13.6	87	0.00
29 T	Tetrahydrofuran	0.108	0.133	-23.1	108	0.00
30 C	Chloroform	0.930	0.992	-6.7#	87	0.00
31 T	Cyclohexane	0.728	0.613	15.8	73	0.00
32 T	1,1,1-Trichloroethane	0.791	0.821	-3.8	83	0.00
33 S	1,2-Dichloroethane-d4	0.517	0.550	-6.4	88	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	79	0.00
35 S	Dibromofluoromethane	0.307	0.323	-5.2	85	0.00
36 T	1,1-Dichloropropene	0.401	0.380	5.2	78	0.00
37 T	Ethyl Acetate	0.227	0.271	-19.4	106	0.00
38 T	Carbon Tetrachloride	0.437	0.432	1.1	81	0.00
39 T	Methylcyclohexane	0.447	0.410	8.3	75	0.00
40 TM	Benzene	1.200	1.209	-0.8	84	0.00
41 T	Methacrylonitrile	0.119	0.131	-10.1	109	0.00
42 TM	1,2-Dichloroethane	0.341	0.369	-8.2	91	0.00
43 T	Isopropyl Acetate	0.404	0.486	-20.3	104	0.00
44 TM	Trichloroethene	0.323	0.320	0.9	82	0.00
45 C	1,2-Dichloropropane	0.316	0.343	-8.5#	89	0.00
46 T	Dibromomethane	0.177	0.189	-6.8	91	0.00
47 T	Bromodichloromethane	0.436	0.481	-10.3	90	0.00
48 T	Methyl methacrylate	0.180	0.216	-20.0	104	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103122\
 Data File : VY011208.D
 Acq On : 31 Oct 2022 19:27
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 01 02:46:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.003	-50.0#	110	0.00
50 S	Toluene-d8	1.218	1.257	-3.2	81	0.00
51 T	4-Methyl-2-Pentanone	0.228	0.290	-27.2#	109	0.00
52 CM	Toluene	0.751	0.778	-3.6#	84	0.00
53 T	t-1,3-Dichloropropene	0.438	0.487	-11.2	92	0.00
54 T	cis-1,3-Dichloropropene	0.496	0.542	-9.3	89	0.00
55 T	1,1,2-Trichloroethane	0.263	0.293	-11.4	95	0.00
56 T	Ethyl methacrylate	0.331	0.403	-21.8	100	0.00
57 T	1,3-Dichloropropane	0.439	0.489	-11.4	94	0.00
58 T	2-Chloroethyl Vinyl ether	0.140	0.192	-37.1#	100	0.00
59 T	2-Hexanone	0.157	0.199	-26.8#	108	0.00
60 T	Dibromochloromethane	0.311	0.351	-12.9	94	0.00
61 T	1,2-Dibromoethane	0.239	0.262	-9.6	94	0.00
62 S	4-Bromofluorobenzene	0.399	0.452	-13.3	91	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	82	0.00
64 T	Tetrachloroethene	0.345	0.337	2.3	83	0.00
65 PM	Chlorobenzene	0.921	0.949	-3.0	86	0.00
66 T	1,1,1,2-Tetrachloroethane	0.357	0.387	-8.4	91	0.00
67 C	Ethyl Benzene	1.608	1.675	-4.2#	85	0.00
68 T	m/p-Xylenes	0.613	0.644	-5.1	86	0.00
69 T	o-Xylene	0.585	0.632	-8.0	87	0.00
70 T	Styrene	0.999	1.113	-11.4	89	0.00
71 P	Bromoform	0.219	0.247	-12.8	97	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	87	0.00
73 T	Isopropylbenzene	3.265	3.333	-2.1	86	0.00
74 T	N-amyl acetate	0.801	0.939	-17.2	105	0.00
75 P	1,1,2,2-Tetrachloroethane	0.712	0.776	-9.0	102	0.00
76 T	1,2,3-Trichloropropane	0.520	0.570	-9.6	104	0.00
77 T	Bromobenzene	0.768	0.781	-1.7	89	0.00
78 T	n-propylbenzene	3.970	4.072	-2.6	87	0.00
79 T	2-Chlorotoluene	2.271	2.346	-3.3	90	0.00
80 T	1,3,5-Trimethylbenzene	2.751	2.851	-3.6	88	0.00
81 T	trans-1,4-Dichloro-2-butene	0.245	0.271	-10.6	102	0.00
82 T	4-Chlorotoluene	2.355	2.423	-2.9	89	0.00
83 T	tert-Butylbenzene	2.401	2.508	-4.5	88	0.00
84 T	1,2,4-Trimethylbenzene	2.707	2.850	-5.3	89	0.00
85 T	sec-Butylbenzene	3.590	3.710	-3.3	87	0.00
86 T	p-Isopropyltoluene	2.949	3.112	-5.5	88	0.00
87 T	1,3-Dichlorobenzene	1.559	1.589	-1.9	91	0.00
88 T	1,4-Dichlorobenzene	1.559	1.571	-0.8	90	0.00
89 T	n-Butylbenzene	2.755	2.900	-5.3	88	0.00
90 T	Hexachloroethane	0.616	0.634	-2.9	91	0.00
91 T	1,2-Dichlorobenzene	1.402	1.461	-4.2	93	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.115	0.132	-14.8	112	0.00
93 T	1,2,4-Trichlorobenzene	0.802	0.883	-10.1	95	0.00
94 T	Hexachlorobutadiene	0.485	0.505	-4.1	90	0.00
95 T	Naphthalene	1.606	1.962	-22.2	107	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103122\
 Data File : VY011208.D
 Acq On : 31 Oct 2022 19:27
 Operator : KP/MD
 Sample : VSTDCCC050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 LabSampleId :
 VSTDCCC050

Quant Time: Nov 01 02:46:29 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102822S.M
 Quant Title : SW846 8260
 QLast Update : Fri Oct 28 23:42:14 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T 1,2,3-Trichlorobenzene	0.699	0.797	-14.0	99	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6