

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103123\
 Data File : VY016148.D
 Acq On : 31 Oct 2023 15:03
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICVVY103123

Quant Time: Nov 01 03:36:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 01 03:33:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	100	0.00
2 T	Dichlorodifluoromethane	0.319	0.325	-1.9	97	0.00
3 P	Chloromethane	0.442	0.422	4.5	97	0.00
4 C	Vinyl Chloride	0.456	0.467	-2.4#	99	0.00
5 T	Bromomethane	0.308	0.330	-7.1	109	0.00
6 T	Chloroethane	0.321	0.322	-0.3	101	0.00
7 T	Trichlorofluoromethane	0.688	0.693	-0.7	98	0.00
8 T	Diethyl Ether	0.228	0.237	-3.9	109	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.433	0.435	-0.5	99	0.00
10 T	Methyl Iodide	0.482	0.518	-7.5	108	0.00
11 T	Tert butyl alcohol	0.035	0.034	2.9	121	0.00
12 CM	1,1-Dichloroethene	0.396	0.399	-0.8#	102	0.00
13 T	Acrolein	0.035	0.033	5.7	120	0.00
14 T	Allyl chloride	0.566	0.586	-3.5	105	0.00
15 T	Acrylonitrile	0.097	0.103	-6.2	114	0.00
16 T	Acetone	0.086	0.078	9.3	113	0.00
17 T	Carbon Disulfide	0.988	1.044	-5.7	103	0.00
18 T	Methyl Acetate	0.376	0.400	-6.4	116	0.00
19 T	Methyl tert-butyl Ether	1.125	1.183	-5.2	110	0.00
20 T	Methylene Chloride	0.533	0.480	9.9	104	0.00
21 T	trans-1,2-Dichloroethene	0.461	0.472	-2.4	103	0.00
22 T	Diisopropyl ether	1.369	1.418	-3.6	107	0.00
23 T	Vinyl Acetate	0.672	0.712	-6.0	111	0.00
24 P	1,1-Dichloroethane	0.819	0.835	-2.0	105	0.00
25 T	2-Butanone	0.129	0.132	-2.3	116	0.00
26 T	2,2-Dichloropropane	0.774	0.794	-2.6	102	0.00
27 T	cis-1,2-Dichloroethene	0.542	0.556	-2.6	105	0.00
28 T	Bromochloromethane	0.307	0.298	2.9	98	0.00
29 T	Tetrahydrofuran	0.081	0.086	-6.2	114	0.00
30 C	Chloroform	0.892	0.913	-2.4#	105	0.00
31 T	Cyclohexane	0.719	0.695	3.3	101	0.00
32 T	1,1,1-Trichloroethane	0.818	0.831	-1.6	102	0.00
33 S	1,2-Dichloroethane-d4	0.468	0.491	-4.9	112	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	102	0.00
35 S	Dibromofluoromethane	0.298	0.308	-3.4	111	0.00
36 T	1,1-Dichloropropene	0.430	0.428	0.5	101	0.00
37 T	Ethyl Acetate	0.192	0.198	-3.1	111	0.00
38 T	Carbon Tetrachloride	0.426	0.449	-5.4	103	0.00
39 T	Methylcyclohexane	0.514	0.522	-1.6	99	0.00
40 TM	Benzene	1.236	1.246	-0.8	105	0.00
41 T	Methacrylonitrile	0.111	0.107	3.6	114	0.00
42 TM	1,2-Dichloroethane	0.345	0.351	-1.7	108	0.00
43 T	Isopropyl Acetate	0.379	0.399	-5.3	115	0.00
44 TM	Trichloroethene	0.362	0.366	-1.1	103	0.00
45 C	1,2-Dichloropropane	0.297	0.307	-3.4#	105	0.00
46 T	Dibromomethane	0.169	0.173	-2.4	107	0.00
47 T	Bromodichloromethane	0.428	0.443	-3.5	108	0.00
48 T	Methyl methacrylate	0.172	0.181	-5.2	113	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103123\
 Data File : VY016148.D
 Acq On : 31 Oct 2023 15:03
 Operator : SY/MD
 Sample : VSTDICV050
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 ICSVY103123

Quant Time: Nov 01 03:36:39 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 01 03:33:29 2023
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.002	0.002	0.0	114	0.00
50 S	Toluene-d8	1.182	1.250	-5.8	110	0.00
51 T	4-Methyl-2-Pentanone	0.189	0.202	-6.9	114	0.00
52 CM	Toluene	0.801	0.817	-2.0#	104	0.00
53 T	t-1,3-Dichloropropene	0.428	0.448	-4.7	110	0.00
54 T	cis-1,3-Dichloropropene	0.502	0.514	-2.4	108	0.00
55 T	1,1,2-Trichloroethane	0.237	0.243	-2.5	108	0.00
56 T	Ethyl methacrylate	0.308	0.335	-8.8	114	0.00
57 T	1,3-Dichloropropane	0.404	0.409	-1.2	109	0.00
58 T	2-Chloroethyl Vinyl ether	0.149	0.151	-1.3	109	0.00
59 T	2-Hexanone	0.130	0.137	-5.4	113	0.00
60 T	Dibromochloromethane	0.298	0.307	-3.0	108	0.00
61 T	1,2-Dibromoethane	0.226	0.233	-3.1	109	0.00
62 S	4-Bromofluorobenzene	0.404	0.416	-3.0	114	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	102	0.00
64 T	Tetrachloroethene	0.445	0.436	2.0	100	0.00
65 PM	Chlorobenzene	0.993	1.012	-1.9	105	0.00
66 T	1,1,1,2-Tetrachloroethane	0.368	0.373	-1.4	105	0.00
67 C	Ethyl Benzene	1.777	1.833	-3.2#	104	0.00
68 T	m/p-Xylenes	0.687	0.705	-2.6	103	0.00
69 T	o-Xylene	0.656	0.668	-1.8	104	0.00
70 T	Styrene	1.090	1.120	-2.8	105	0.00
71 P	Bromoform	0.204	0.212	-3.9	111	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	100	0.00
73 T	Isopropylbenzene	3.751	3.856	-2.8	103	0.00
74 T	N-amyl acetate	0.815	0.886	-8.7	113	0.00
75 P	1,1,2,2-Tetrachloroethane	0.597	0.630	-5.5	115	0.00
76 T	1,2,3-Trichloropropane	0.471	0.492	-4.5	99	0.00
77 T	Bromobenzene	0.849	0.886	-4.4	108	0.00
78 T	n-propylbenzene	4.434	4.566	-3.0	102	0.00
79 T	2-Chlorotoluene	2.473	2.521	-1.9	104	0.00
80 T	1,3,5-Trimethylbenzene	3.039	3.119	-2.6	103	0.00
81 T	trans-1,4-Dichloro-2-butene	0.217	0.235	-8.3	115	0.00
82 T	4-Chlorotoluene	2.540	2.630	-3.5	106	0.00
83 T	tert-Butylbenzene	2.719	2.797	-2.9	102	0.00
84 T	1,2,4-Trimethylbenzene	3.007	3.079	-2.4	104	0.00
85 T	sec-Butylbenzene	3.991	4.117	-3.2	103	0.00
86 T	p-Isopropyltoluene	3.307	3.434	-3.8	103	0.00
87 T	1,3-Dichlorobenzene	1.680	1.723	-2.6	105	0.00
88 T	1,4-Dichlorobenzene	1.660	1.677	-1.0	106	0.00
89 T	n-Butylbenzene	3.039	3.115	-2.5	102	0.00
90 T	Hexachloroethane	0.570	0.594	-4.2	104	0.00
91 T	1,2-Dichlorobenzene	1.451	1.489	-2.6	106	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.107	0.105	1.9	106	0.00
93 T	1,2,4-Trichlorobenzene	0.896	0.930	-3.8	106	0.00
94 T	Hexachlorobutadiene	0.489	0.497	-1.6	102	0.00
95 T	Naphthalene	1.740	1.814	-4.3	111	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY103123\
Data File : VY016148.D
Acq On : 31 Oct 2023 15:03
Operator : SY/MD
Sample : VSTDICV050
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
ICVVY103123

Quant Time: Nov 01 03:36:39 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 01 03:33:29 2023
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.765	0.794	-3.8	108	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6