

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110121\
 Data File : VY006552.D
 Acq On : 01 Nov 2021 15:20
 Operator : SY/MD
 Sample : M4429-01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 OK-03-102921

Quant Time: Nov 02 03:29:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	5123	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	8067	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.489	117	7617	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.422	152	2552	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.142	65	3617	57.375	ug/l	0.00
Spiked Amount	50.000	Range 50 - 163	Recovery	=	114.740%	
35) Dibromofluoromethane	7.722	113	836	18.593	ug/l	0.00
Spiked Amount	50.000	Range 54 - 147	Recovery	=	37.180%#	
50) Toluene-d8	10.179	98	10402	54.573	ug/l	0.00
Spiked Amount	50.000	Range 49 - 140	Recovery	=	109.140%	
62) 4-Bromofluorobenzene	12.483	95	3328	51.832	ug/l	0.00
Spiked Amount	50.000	Range 25 - 144	Recovery	=	103.660%	
Target Compounds						
					Qvalue	
3) Chloromethane	2.119	50	285	4.459	ug/l #	43
5) Bromomethane	2.668	94	226	4.752	ug/l #	72
10) Methyl Iodide	4.107	142	663	16.869	ug/l #	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110121\
 Data File : VY006552.D
 Acq On : 01 Nov 2021 15:20
 Operator : SY/MD
 Sample : M4429-01
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_Y
ClientSampleId :
 OK-03-102921

Quant Time: Nov 02 03:29:58 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

