

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110121\
 Data File : VY006566.D
 Acq On : 01 Nov 2021 20:50
 Operator : SY/MD
 Sample : M4428-01
 Misc : 5.05G/5ML/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 HD-01-102921RE

Quant Time: Nov 02 03:32:24 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	3745	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.691	114	5966	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	6026	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	2267	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	3371	73.148	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	146.300%
35) Dibromofluoromethane	7.722	113	439	13.202	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	26.400%#
50) Toluene-d8	10.185	98	7720	54.766	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	109.540%
62) 4-Bromofluorobenzene	12.483	95	2396	50.458	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	100.920%
Target Compounds						Qvalue
5) Bromomethane	2.674	94	274	7.881	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110121\
Data File : VY006566.D
Acq On : 01 Nov 2021 20:50
Operator : SY/MD
Sample : M4428-01
Misc : 5.05G/5ML/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
HD-01-102921RE

Quant Time: Nov 02 03:32:24 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

