

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110123\
 Data File : VY016163.D
 Acq On : 01 Nov 2023 13:06
 Operator : SY/MD
 Sample : BLK
 Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 BLK

Quant Time: Nov 02 02:23:46 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 01 03:33:29 2023
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	124155	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	209056	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.502	117	192703	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	90369	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	67429	58.040	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	= 116.080%	
35) Dibromofluoromethane	7.728	113	57543	46.170	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	= 92.340%	
50) Toluene-d8	10.191	98	243582	49.294	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	= 98.580%	
62) 4-Bromofluorobenzene	12.489	95	85784	50.768	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	= 101.540%	
Target Compounds						Qvalue
20) Methylene Chloride	4.728	84	12507	8.028	ug/l #	80
30) Chloroform	7.521	83	7257	3.277	ug/l	98
47) Bromodichloromethane	9.508	83	3353	1.872	ug/l #	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110123\
Data File : VY016163.D
Acq On : 01 Nov 2023 13:06
Operator : SY/MD
Sample : BLK
Misc : 5.00g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
BLK

Quant Time: Nov 02 02:23:46 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 01 03:33:29 2023
Response via : Initial Calibration

