

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110221\
 Data File : VY006581.D
 Acq On : 02 Nov 2021 17:01
 Operator : SY/MD
 Sample : M4424-10
 Misc : 5.00G/5ML/MSVOA_Y/SOIL
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 T3-EP02-NE

Quant Time: Nov 03 02:02:16 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.795	168	12122	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	17528	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	12375	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3579	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	5768	38.668	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	77.340%
35) Dibromofluoromethane	7.728	113	1562	15.988	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	31.980%#
50) Toluene-d8	10.185	98	20210	48.799	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.600%
62) 4-Bromofluorobenzene	12.483	95	4585	32.865	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	65.740%
Target Compounds						
83) tert-Butylbenzene	13.081	119	479	2.659	ug/l	Qvalue 95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110221\
Data File : VY006581.D
Acq On : 02 Nov 2021 17:01
Operator : SY/MD
Sample : M4424-10
Misc : 5.00G/5ML/MSVOA_Y/SOIL
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
T3-EP02-NE

Quant Time: Nov 03 02:02:16 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

