

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110223\
 Data File : VY016190.D
 Acq On : 02 Nov 2023 15:11
 Operator : SY/MD
 Sample : 05214-05
 Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 NL-OILY-DEBRIS

Quant Time: Nov 03 01:41:27 2023
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
 Quant Title : SW846 8260
 QLast Update : Wed Nov 01 03:33:29 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	123719	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.703	114	206828	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	196381	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.434	152	75825	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	67160	58.013	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	116.020%
35) Dibromofluoromethane	7.728	113	56271	45.636	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	91.280%
50) Toluene-d8	10.191	98	249829	51.102	ug/l	0.00
Spiked Amount	50.000	Range	58 - 134	Recovery	=	102.200%
62) 4-Bromofluorobenzene	12.489	95	83375	49.874	ug/l	0.00
Spiked Amount	50.000	Range	30 - 143	Recovery	=	99.740%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	4.978	59	31516	364.364	ug/l #	82
16) Acetone	3.966	43	9331	34.007	ug/l	100
20) Methylene Chloride	4.734	84	13308	8.792	ug/l #	88
42) 1,2-Dichloroethane	8.252	62	2970	2.082	ug/l	98
52) Toluene	10.252	92	9881	2.983	ug/l	94

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110223\
Data File : VY016190.D
Acq On : 02 Nov 2023 15:11
Operator : SY/MD
Sample : 05214-05
Misc : 5.05g/5.0mL/MSVOA_Y/SOIL
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
NL-OILY-DEBRIS

Quant Time: Nov 03 01:41:27 2023
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y103123S.M
Quant Title : SW846 8260
QLast Update : Wed Nov 01 03:33:29 2023
Response via : Initial Calibration

