

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110321\
 Data File : VY006598.D
 Acq On : 03 Nov 2021 13:33
 Operator : SY/MD
 Sample : M4433-03RE
 Misc : 4.72G/5ML/MSVOA_Y/SOIL
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 PSC-124035RE

Quant Time: Nov 03 21:19:09 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	66782	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	107035	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	85748	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	30755	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.149	65	37368	45.471	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	90.940%
35) Dibromofluoromethane	7.728	113	31221	52.333	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	104.660%
50) Toluene-d8	10.185	98	131885	52.149	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	104.300%
62) 4-Bromofluorobenzene	12.483	95	34182	40.124	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	80.240%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110321\
Data File : VY006598.D
Acq On : 03 Nov 2021 13:33
Operator : SY/MD
Sample : M4433-03RE
Misc : 4.72G/5ML/MSVOA_Y/SOIL
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
PSC-124035RE

Quant Time: Nov 03 21:19:09 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

