

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110321\
 Data File : VY006615.D
 Acq On : 03 Nov 2021 20:39
 Operator : SY/MD
 Sample : M4436-01
 Misc : 5.01G/5ML/MSVOA_Y/SOIL
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 S15-SP

Quant Time: Nov 03 21:24:51 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	24281	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	34910	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.496	117	19668	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	3347	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.155	65	9072	30.362	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	60.720%
35) Dibromofluoromethane	7.728	113	3896	20.023	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	40.040%#
50) Toluene-d8	10.185	98	40001	48.495	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	97.000%
62) 4-Bromofluorobenzene	12.483	95	5049	18.171	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	36.340%
Target Compounds						Qvalue
83) tert-Butylbenzene	13.081	119	963	5.716	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110321\
Data File : VY006615.D
Acq On : 03 Nov 2021 20:39
Operator : SY/MD
Sample : M4436-01
Misc : 5.01G/5ML/MSVOA_Y/SOIL
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
S15-SP

Quant Time: Nov 03 21:24:51 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

