

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110321\
 Data File : VY006616.D
 Acq On : 03 Nov 2021 21:03
 Operator : SY/MD
 Sample : M4438-01
 Misc : 4.46G/5ML/MSVOA_Y/SOIL
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_Y
 ClientSampleId :
 FDR-20211101-WC-S

Quant Time: Nov 03 21:59:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
 Quant Title : SW846 8260
 QLast Update : Sat Oct 23 06:59:04 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.801	168	138721	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.697	114	233969	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.495	117	201742	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.428	152	68033	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.148	65	103654	60.721	ug/l	0.00
Spiked Amount	50.000	Range	50 - 163	Recovery	=	121.440%
35) Dibromofluoromethane	7.728	113	72003	55.214	ug/l	0.00
Spiked Amount	50.000	Range	54 - 147	Recovery	=	110.420%
50) Toluene-d8	10.184	98	295077	53.377	ug/l	0.00
Spiked Amount	50.000	Range	49 - 140	Recovery	=	106.760%
62) 4-Bromofluorobenzene	12.483	95	80891	43.438	ug/l	0.00
Spiked Amount	50.000	Range	25 - 144	Recovery	=	86.880%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_Y\Data\VY110321\
Data File : VY006616.D
Acq On : 03 Nov 2021 21:03
Operator : SY/MD
Sample : M4438-01
Misc : 4.46G/5ML/MSVOA_Y/SOIL
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_Y
ClientSampleId :
FDR-20211101-WC-S

Quant Time: Nov 03 21:59:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_Y\methods\82Y102221S.M
Quant Title : SW846 8260
QLast Update : Sat Oct 23 06:59:04 2021
Response via : Initial Calibration

